

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN010521\
 Data File : VN065378.D
 Acq On : 5 Jan 2021 10:38
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Jan 06 01:13:41 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122220W.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 22 15:42:18 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	123	0.00
2 T	Dichlorodifluoromethane	0.463	0.437	5.6	109	0.00
3 P	Chloromethane	0.527	0.539	-2.3	121	0.00
4 C	Vinyl Chloride	0.548	0.574	-4.7#	123	0.00
5 T	Bromomethane	0.389	0.409	-5.1	127	0.00
6 T	Chloroethane	0.354	0.374	-5.6	124	0.00
7 T	Trichlorofluoromethane	0.820	0.778	5.1	113	0.00
8 T	Diethyl Ether	0.323	0.316	2.2	121	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.460	0.457	0.7	121	0.00
10 T	Methyl Iodide	0.694	0.682	1.7	114	0.00
11 T	Tert butyl alcohol	0.092	0.075	18.5	97	0.00
12 CM	1,1-Dichloroethene	0.491	0.457	6.9#	116	0.00
13 T	Acrolein	0.065	0.063	3.1	121	0.00
14 T	Allyl chloride	0.810	0.737	9.0	109	0.00
15 T	Acrylonitrile	0.223	0.218	2.2	115	0.00
16 T	Acetone	0.260	0.241	7.3	122	0.00
17 T	Carbon Disulfide	1.257	1.144	9.0	107	0.00
18 T	Methyl Acetate	0.507	0.501	1.2	118	0.00
19 T	Methyl tert-butyl Ether	1.560	1.465	6.1	114	0.00
20 T	Methylene Chloride	0.551	0.513	6.9	122	0.00
21 T	trans-1,2-Dichloroethene	0.482	0.481	0.2	120	0.00
22 T	Diisopropyl ether	1.611	1.545	4.1	117	0.00
23 T	Vinyl Acetate	1.358	1.255	7.6	109	0.00
24 P	1,1-Dichloroethane	0.884	0.882	0.2	119	0.00
25 T	2-Butanone	0.317	0.305	3.8	114	0.00
26 T	2,2-Dichloropropane	0.844	0.773	8.4	113	0.00
27 T	cis-1,2-Dichloroethene	0.561	0.576	-2.7	123	0.00
28 T	Bromochloromethane	0.379	0.366	3.4	119	0.00
29 T	Tetrahydrofuran	0.205	0.186	9.3	110	0.00
30 C	Chloroform	0.910	0.909	0.1#	120	0.00
31 T	Cyclohexane	0.833	0.753	9.6	114	0.00
32 T	1,1,1-Trichloroethane	0.824	0.779	5.5	114	0.00
33 S	1,2-Dichloroethane-d4	0.564	0.524	7.1	113	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	120	0.00
35 S	Dibromofluoromethane	0.302	0.314	-4.0	123	0.00
36 T	1,1-Dichloropropene	0.455	0.464	-2.0	121	0.00
37 T	Ethyl Acetate	0.451	0.406	10.0	105	0.00
38 T	Carbon Tetrachloride	0.477	0.464	2.7	115	0.00
39 T	Methylcyclohexane	0.546	0.550	-0.7	116	0.00
40 TM	Benzene	1.352	1.414	-4.6	124	0.00
41 T	Methacrylonitrile	0.234	0.204	12.8	108	0.00
42 TM	1,2-Dichloroethane	0.500	0.469	6.2	112	0.00
43 T	Isopropyl Acetate	0.777	0.705	9.3	105	0.00
44 TM	Trichloroethene	0.390	0.409	-4.9	124	0.00
45 C	1,2-Dichloropropane	0.352	0.370	-5.1#	123	0.00
46 T	Dibromomethane	0.239	0.246	-2.9	122	0.00
47 T	Bromodichloromethane	0.503	0.503	0.0	117	0.00
48 T	Methyl methacrylate	0.377	0.347	8.0	108	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN010521\
 Data File : VN065378.D
 Acq On : 5 Jan 2021 10:38
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Jan 06 01:13:41 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122220W.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 22 15:42:18 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.006	0.006	0.0	118	0.00
50 S	Toluene-d8	1.131	1.195	-5.7	126	0.00
51 T	4-Methyl-2-Pentanone	0.437	0.415	5.0	110	0.00
52 CM	Toluene	0.852	0.913	-7.2#	126	0.00
53 T	t-1,3-Dichloropropene	0.567	0.570	-0.5	118	0.00
54 T	cis-1,3-Dichloropropene	0.603	0.610	-1.2	120	0.00
55 T	1,1,2-Trichloroethane	0.337	0.356	-5.6	124	0.00
56 T	Ethyl methacrylate	0.522	0.530	-1.5	119	0.00
57 T	1,3-Dichloropropane	0.557	0.595	-6.8	124	0.00
58 T	2-Chloroethyl Vinyl ether	0.227	0.224	1.3	118	0.00
59 T	2-Hexanone	0.321	0.314	2.2	112	0.00
60 T	Dibromochloromethane	0.402	0.413	-2.7	120	0.00
61 T	1,2-Dibromoethane	0.362	0.378	-4.4	122	0.00
62 S	4-Bromofluorobenzene	0.424	0.444	-4.7	125	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	122	0.00
64 T	Tetrachloroethene	0.486	0.491	-1.0	126	0.00
65 PM	Chlorobenzene	1.018	1.086	-6.7	128	0.00
66 T	1,1,1,2-Tetrachloroethane	0.396	0.414	-4.5	125	0.00
67 C	Ethyl Benzene	1.834	1.903	-3.8#	125	0.00
68 T	m/p-Xylenes	0.691	0.733	-6.1	128	0.00
69 T	o-Xylene	0.671	0.713	-6.3	128	0.00
70 T	Styrene	1.119	1.216	-8.7	129	0.00
71 P	Bromoform	0.341	0.345	-1.2	119	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	127	0.00
73 T	Isopropylbenzene	3.808	3.681	3.3	127	0.00
74 T	N-amyl acetate	1.516	1.328	12.4	109	0.00
75 P	1,1,2,2-Tetrachloroethane	1.027	1.004	2.2	127	0.00
76 T	1,2,3-Trichloropropane	1.020	1.037	-1.7	138	0.00
77 T	Bromobenzene	0.995	1.001	-0.6	131	0.00
78 T	n-propylbenzene	4.299	4.166	3.1	126	0.00
79 T	2-Chlorotoluene	2.532	2.456	3.0	125	0.00
80 T	1,3,5-Trimethylbenzene	3.247	3.129	3.6	126	0.00
81 T	trans-1,4-Dichloro-2-butene	0.405	0.362	10.6	113	0.00
82 T	4-Chlorotoluene	2.607	2.555	2.0	127	0.00
83 T	tert-Butylbenzene	2.767	2.692	2.7	127	0.00
84 T	1,2,4-Trimethylbenzene	3.162	3.154	0.3	129	0.00
85 T	sec-Butylbenzene	3.671	3.546	3.4	126	0.00
86 T	p-Isopropyltoluene	3.365	3.275	2.7	126	0.00
87 T	1,3-Dichlorobenzene	1.664	1.739	-4.5	133	0.00
88 T	1,4-Dichlorobenzene	1.671	1.715	-2.6	131	0.00
89 T	n-Butylbenzene	2.876	2.769	3.7	122	0.00
90 T	Hexachloroethane	0.579	0.565	2.4	122	0.00
91 T	1,2-Dichlorobenzene	1.598	1.699	-6.3	133	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.212	0.195	8.0	113	0.00
93 T	1,2,4-Trichlorobenzene	0.806	0.978	-21.3	132	0.00
94 T	Hexachlorobutadiene	0.631	0.597	5.4	128	0.00
95 T	Naphthalene	2.105	2.425	-15.2	128	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN010521\
Data File : VN065378.D
Acq On : 5 Jan 2021 10:38
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: Jan 06 01:13:41 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122220W.M
Quant Title : SW846 8260
QLast Update : Tue Dec 22 15:42:18 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.764	0.910	-19.1	135	0.02

(#) = Out of Range SPCC's out = 0 CCC's out = 6