

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN010522\
 Data File : VN070487.D
 Acq On : 5 Jan 2022 14:15
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 IBLK

Quant Time: Jan 06 12:34:33 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122721W.M
 Quant Title : SW846 8260
 QLast Update : Mon Dec 27 18:47:12 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	1138139	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.968	114	2193634	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.744	117	2648439	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.675	152	1101770	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.437	65	810019	49.499	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	99.000%
35) Dibromofluoromethane	8.021	113	623904	47.007	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	94.020%
50) Toluene-d8	10.440	98	2948695	54.384	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	108.760%
62) 4-Bromofluorobenzene	12.731	95	1311361	66.964	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	133.920%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN010522\
Data File : VN070487.D
Acq On : 5 Jan 2022 14:15
Operator : JC/MD
Sample : IBLK
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
IBLK

Quant Time: Jan 06 12:34:33 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122721W.M
Quant Title : SW846 8260
QLast Update : Mon Dec 27 18:47:12 2021
Response via : Initial Calibration

