

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN010522\
 Data File : VN070488.D
 Acq On : 5 Jan 2022 14:40
 Operator : JC/MD
 Sample : N1012-02DL 40X
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 31212DL

Quant Time: Jan 06 12:34:42 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122721W.M
 Quant Title : SW846 8260
 QLast Update : Mon Dec 27 18:47:12 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.083	168	1796181	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.965	114	2998567	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.744	117	2679911	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.675	152	941512	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.437	65	1470627	56.944	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	113.880%
35) Dibromofluoromethane	8.021	113	1022469	56.357	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	112.720%
50) Toluene-d8	10.440	98	4235368	57.146	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	114.300%
62) 4-Bromofluorobenzene	12.731	95	1536005	57.380	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	114.760%
Target Compounds						Qvalue
16) Acetone	4.291	43	24517	1.462	ug/l	99
25) 2-Butanone	7.335	43	625346	29.625	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN010522\
Data File : VN070488.D
Acq On : 5 Jan 2022 14:40
Operator : JC/MD
Sample : N1012-02DL 40X
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
31212DL

Quant Time: Jan 06 12:34:42 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122721W.M
Quant Title : SW846 8260
QLast Update : Mon Dec 27 18:47:12 2021
Response via : Initial Calibration

