

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN010522\
 Data File : VN070501.D
 Acq On : 5 Jan 2022 20:10
 Operator : JC/MD
 Sample : N1015-02 100X
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 30915

Quant Time: Jan 06 13:27:23 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122721W.M
 Quant Title : SW846 8260
 QLast Update : Mon Dec 27 18:47:12 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	609570	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.968	114	1022880	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.744	117	909228	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.675	152	292848	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.437	65	663687	75.725	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 151.440%#	
35) Dibromofluoromethane	8.021	113	468677	75.728	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 151.460%#	
50) Toluene-d8	10.440	98	1866961	73.844	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 147.680%#	
62) 4-Bromofluorobenzene	12.731	95	635994	69.648	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 139.300%#	
Target Compounds						
						Qvalue
16) Acetone	4.299	43	27649	4.860	ug/l	93
20) Methylene Chloride	5.103	84	7337	0.932	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN010522\
Data File : VN070501.D
Acq On : 5 Jan 2022 20:10
Operator : JC/MD
Sample : N1015-02 100X
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
30915

Quant Time: Jan 06 13:27:23 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122721W.M
Quant Title : SW846 8260
QLast Update : Mon Dec 27 18:47:12 2021
Response via : Initial Calibration

