

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN010522\
 Data File : VN070503.D
 Acq On : 5 Jan 2022 21:01
 Operator : JC/MD
 Sample : N1015-04 100X
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 20969

Quant Time: Jan 06 13:27:41 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122721W.M
 Quant Title : SW846 8260
 QLast Update : Mon Dec 27 18:47:12 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	880239	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.968	114	1499563	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.744	117	1331963	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.675	152	433825	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.437	65	734690	58.050	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	116.100%
35) Dibromofluoromethane	8.021	113	517487	57.035	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	114.080%
50) Toluene-d8	10.440	98	2006037	54.123	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	108.240%
62) 4-Bromofluorobenzene	12.731	95	706281	52.759	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	105.520%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN010522\
Data File : VN070503.D
Acq On : 5 Jan 2022 21:01
Operator : JC/MD
Sample : N1015-04 100X
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
20969

Quant Time: Jan 06 13:27:41 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122721W.M
Quant Title : SW846 8260
QLast Update : Mon Dec 27 18:47:12 2021
Response via : Initial Calibration

