

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN010523\
 Data File : VN076104.D
 Acq On : 05 Jan 2023 11:36
 Operator : JC\MD
 Sample : VN0105WBL01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0105WBL01

Quant Time: Jan 05 23:46:54 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121422W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 14 13:03:41 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	329623	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	513556	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	436310	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	177867	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	171642	47.865	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	95.720%
35) Dibromofluoromethane	8.177	113	154348	48.531	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.060%
50) Toluene-d8	10.571	98	587907	49.441	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	98.880%
62) 4-Bromofluorobenzene	12.853	95	177516	45.073	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	90.140%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN010523\
Data File : VN076104.D
Acq On : 05 Jan 2023 11:36
Operator : JC\MD
Sample : VN0105WBL01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0105WBL01

Quant Time: Jan 05 23:46:54 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121422W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 14 13:03:41 2022
Response via : Initial Calibration

