

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN010523\
 Data File : VN076112.D
 Acq On : 05 Jan 2023 15:39
 Operator : JC\MD
 Sample : 01003-01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 STORAGE-BLANK-SOIL-REF-6

Quant Time: Jan 05 23:50:06 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121422W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 14 13:03:41 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.235	168	330220	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	513195	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	445858	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	182525	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	170686	47.512	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	95.020%
35) Dibromofluoromethane	8.177	113	155544	48.942	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.880%
50) Toluene-d8	10.571	98	593837	49.975	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	99.940%
62) 4-Bromofluorobenzene	12.853	95	186636	47.421	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	94.840%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN010523\
Data File : VN076112.D
Acq On : 05 Jan 2023 15:39
Operator : JC\MD
Sample : 01003-01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
STORAGE-BLANK-SOIL-REF-6

Quant Time: Jan 05 23:50:06 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121422W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 14 13:03:41 2022
Response via : Initial Calibration

