

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN010621\  
 Data File : VN065402.D  
 Acq On : 6 Jan 2021 16:03  
 Operator : JC/MD  
 Sample : M1008-17  
 Misc : 5.00mL/MSVOA\_N/WATER  
 ALS Vial : 16 Sample Multiplier: 1

Instrument :  
 MSVOA\_N  
 ClientSampleId :  
 375-393

Quant Time: Jan 07 01:28:42 2021  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N122220W.M  
 Quant Title : SW846 8260  
 QLast Update : Tue Dec 22 15:42:18 2020  
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
-----						
Internal Standards						
1) Pentafluorobenzene	7.630	168	258313	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.553	114	428533	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.380	117	465820	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.318	152	204464	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.991	65	146969	50.40	ug/l	0.00
Spiked Amount	50.000	Range 61 - 141	Recovery	=	100.80%	
35) Dibromofluoromethane	7.553	113	132467	51.24	ug/l	0.00
Spiked Amount	50.000	Range 69 - 133	Recovery	=	102.48%	
50) Toluene-d8	10.061	98	542583	55.99	ug/l	0.00
Spiked Amount	50.000	Range 65 - 126	Recovery	=	111.98%	
62) 4-Bromofluorobenzene	12.376	95	214048	55.26	ug/l	0.00
Spiked Amount	50.000	Range 58 - 135	Recovery	=	110.52%	
Target Compounds						
						Qvalue
16) Acetone	3.791	43	21555	7.16	ug/l	92
20) Methylene Chloride	4.515	84	10016	3.52	ug/l	96
43) Isopropyl Acetate	8.135	43	20762	3.12	ug/l	96
52) Toluene	10.126	92	45460	6.22	ug/l	96
68) m/p-Xylenes	11.592	106	15534	2.41	ug/l	99
69) o-Xylene	11.923	106	9283	1.49	ug/l	98
84) 1,2,4-Trimethylbenzene	13.013	105	20040	1.55	ug/l	97
95) Naphthalene	15.106	128	5352	3.84	ug/l	96
-----						

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN010621\  
 Data File : VN065402.D  
 Acq On : 6 Jan 2021 16:03  
 Operator : JC/MD  
 Sample : M1008-17  
 Misc : 5.00mL/MSVOA\_N/WATER  
 ALS Vial : 16 Sample Multiplier: 1

Instrument :  
 MSVOA\_N  
 ClientSampleId :  
 375-393

Quant Time: Jan 07 01:28:42 2021  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N122220W.M  
 Quant Title : SW846 8260  
 QLast Update : Tue Dec 22 15:42:18 2020  
 Response via : Initial Calibration

