

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN010621\
 Data File : VN065403.D
 Acq On : 6 Jan 2021 16:28
 Operator : JC/MD
 Sample : M1022-05 50X
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 41098

Quant Time: Jan 07 01:29:01 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122220W.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 22 15:42:18 2020
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.630	168	255531	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.553	114	427939	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.379	117	457159	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.318	152	204945	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.991	65	146704	50.85	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	101.70%
35) Dibromofluoromethane	7.553	113	132569	51.35	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	102.70%
50) Toluene-d8	10.061	98	534963	55.28	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	110.56%
62) 4-Bromofluorobenzene	12.376	95	215599	55.69	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	111.38%
Target Compounds						
					Qvalue	
16) Acetone	3.791	43	576382	486.31	ug/l	95
18) Methyl Acetate	4.293	43	3207	1.24	ug/l #	89
37) Ethyl Acetate	6.904	43	18810	4.88	ug/l	98
52) Toluene	10.126	92	20579	2.82	ug/l	97
84) 1,2,4-Trimethylbenzene	13.019	105	3665	0.28	ug/l	72
86) p-Isopropyltoluene	13.260	119	45200	3.28	ug/l #	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN010621\
Data File : VN065403.D
Acq On : 6 Jan 2021 16:28
Operator : JC/MD
Sample : M1022-05 50X
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
41098

Quant Time: Jan 07 01:29:01 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122220W.M
Quant Title : SW846 8260
QLast Update : Tue Dec 22 15:42:18 2020
Response via : Initial Calibration

