

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN010621\
 Data File : VN065404.D
 Acq On : 6 Jan 2021 16:53
 Operator : JC/MD
 Sample : M1022-06 50X
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 41099

Quant Time: Jan 07 01:29:18 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122220W.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 22 15:42:18 2020
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.630	168	258211	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.553	114	430816	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.379	117	459006	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.318	152	200499	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.991	65	145782	50.01	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 100.02%	
35) Dibromofluoromethane	7.553	113	130786	50.32	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 100.64%	
50) Toluene-d8	10.061	98	539342	55.37	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 110.74%	
62) 4-Bromofluorobenzene	12.376	95	211193	54.33	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 108.66%	
Target Compounds						
					Qvalue	
16) Acetone	3.791	43	528029	439.84	ug/l	95
18) Methyl Acetate	4.296	43	4065	1.55	ug/l #	89
37) Ethyl Acetate	6.897	43	16688	4.30	ug/l	99
52) Toluene	10.126	92	13348	1.82	ug/l	96
84) 1,2,4-Trimethylbenzene	13.016	105	3329	0.26	ug/l	69
86) p-Isopropyltoluene	13.260	119	34330	2.54	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN010621\
Data File : VN065404.D
Acq On : 6 Jan 2021 16:53
Operator : JC/MD
Sample : M1022-06 50X
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
41099

Quant Time: Jan 07 01:29:18 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122220W.M
Quant Title : SW846 8260
QLast Update : Tue Dec 22 15:42:18 2020
Response via : Initial Calibration

