

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN010621\
 Data File : VN065406.D
 Acq On : 6 Jan 2021 17:42
 Operator : JC/MD
 Sample : M1022-08 50X
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 16790DL

Quant Time: Jan 07 01:29:48 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122220W.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 22 15:42:18 2020
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.630	168	259560	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.553	114	433871	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.379	117	461034	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.318	152	178278	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.990	65	147667	50.39	ug/l	0.00
Spiked Amount	50.000	Range 61 - 141	Recovery	= 100.78%		
35) Dibromofluoromethane	7.553	113	132379	50.57	ug/l	0.00
Spiked Amount	50.000	Range 69 - 133	Recovery	= 101.14%		
50) Toluene-d8	10.061	98	538776	54.92	ug/l	0.00
Spiked Amount	50.000	Range 65 - 126	Recovery	= 109.84%		
62) 4-Bromofluorobenzene	12.376	95	204295	52.38	ug/l	0.00
Spiked Amount	50.000	Range 58 - 135	Recovery	= 104.76%		
Target Compounds						Qvalue
16) Acetone	3.791	43	33597	17.31	ug/l	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN010621\
Data File : VN065406.D
Acq On : 6 Jan 2021 17:42
Operator : JC/MD
Sample : M1022-08 50X
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
16790DL

Quant Time: Jan 07 01:29:48 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122220W.M
Quant Title : SW846 8260
QLast Update : Tue Dec 22 15:42:18 2020
Response via : Initial Calibration

