

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN010621\
 Data File : VN065407.D
 Acq On : 6 Jan 2021 18:07
 Operator : JC/MD
 Sample : M1022-09 50X
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 30734

Quant Time: Jan 07 01:30:01 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122220W.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 22 15:42:18 2020
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.630	168	268409	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.553	114	456538	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.379	117	484602	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.318	152	197647	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.991	65	153891	50.79	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	101.58%
35) Dibromofluoromethane	7.556	113	135382	49.15	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	98.30%
50) Toluene-d8	10.058	98	560083	54.26	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	108.52%
62) 4-Bromofluorobenzene	12.376	95	217271	52.89	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	105.78%
Target Compounds						
16) Acetone	3.791	43	452813	360.88	ug/l	95
52) Toluene	10.126	92	2409	0.31	ug/l	90
64) Tetrachloroethene	10.605	164	4558	0.97	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN010621\
Data File : VN065407.D
Acq On : 6 Jan 2021 18:07
Operator : JC/MD
Sample : M1022-09 50X
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
30734

Quant Time: Jan 07 01:30:01 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122220W.M
Quant Title : SW846 8260
QLast Update : Tue Dec 22 15:42:18 2020
Response via : Initial Calibration

