

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN010621\
 Data File : VN065412.D
 Acq On : 6 Jan 2021 20:10
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Jan 07 01:31:19 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122220W.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 22 15:42:18 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	114	0.00
2 T	Dichlorodifluoromethane	0.463	0.400	13.6	93	0.00
3 P	Chloromethane	0.527	0.499	5.3	104	0.00
4 C	Vinyl Chloride	0.548	0.541	1.3#	107	0.00
5 T	Bromomethane	0.389	0.372	4.4	107	0.00
6 T	Chloroethane	0.354	0.357	-0.8	110	0.00
7 T	Trichlorofluoromethane	0.820	0.746	9.0	100	0.00
8 T	Diethyl Ether	0.323	0.309	4.3	110	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.460	0.444	3.5	109	0.00
10 T	Methyl Iodide	0.694	0.696	-0.3	108	0.00
11 T	Tert butyl alcohol	0.092	0.076	17.4	91	0.00
12 CM	1,1-Dichloroethene	0.491	0.453	7.7#	106	0.00
13 T	Acrolein	0.065	0.060	7.7	108	0.00
14 T	Allyl chloride	0.810	0.701	13.5	96	0.00
15 T	Acrylonitrile	0.223	0.223	0.0	109	0.00
16 T	Acetone	0.260	0.181	30.4#	85	0.00
17 T	Carbon Disulfide	1.257	1.064	15.4	92	0.00
18 T	Methyl Acetate	0.507	0.514	-1.4	112	0.00
19 T	Methyl tert-butyl Ether	1.560	1.449	7.1	104	0.00
20 T	Methylene Chloride	0.551	0.509	7.6	112	0.00
21 T	trans-1,2-Dichloroethene	0.482	0.467	3.1	108	0.00
22 T	Diisopropyl ether	1.611	1.530	5.0	108	0.00
23 T	Vinyl Acetate	1.358	1.247	8.2	100	0.00
24 P	1,1-Dichloroethane	0.884	0.881	0.3	110	0.00
25 T	2-Butanone	0.317	0.282	11.0	98	0.00
26 T	2,2-Dichloropropane	0.844	0.684	19.0	92	0.00
27 T	cis-1,2-Dichloroethene	0.561	0.569	-1.4	113	0.00
28 T	Bromochloromethane	0.379	0.366	3.4	111	0.00
29 T	Tetrahydrofuran	0.205	0.190	7.3	104	0.00
30 C	Chloroform	0.910	0.908	0.2#	111	0.00
31 T	Cyclohexane	0.833	0.732	12.1	103	0.00
32 T	1,1,1-Trichloroethane	0.824	0.768	6.8	104	0.00
33 S	1,2-Dichloroethane-d4	0.564	0.540	4.3	108	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	115	0.00
35 S	Dibromofluoromethane	0.302	0.311	-3.0	117	0.00
36 T	1,1-Dichloropropene	0.455	0.432	5.1	108	0.00
37 T	Ethyl Acetate	0.451	0.405	10.2	101	0.00
38 T	Carbon Tetrachloride	0.477	0.423	11.3	101	0.00
39 T	Methylcyclohexane	0.546	0.493	9.7	100	0.00
40 TM	Benzene	1.352	1.351	0.1	114	0.00
41 T	Methacrylonitrile	0.234	0.202	13.7	103	0.00
42 TM	1,2-Dichloroethane	0.500	0.455	9.0	105	0.00
43 T	Isopropyl Acetate	0.777	0.681	12.4	98	0.00
44 TM	Trichloroethene	0.390	0.389	0.3	113	0.00
45 C	1,2-Dichloropropane	0.352	0.354	-0.6#	114	0.00
46 T	Dibromomethane	0.239	0.232	2.9	111	0.00
47 T	Bromodichloromethane	0.503	0.474	5.8	107	0.00
48 T	Methyl methacrylate	0.377	0.329	12.7	99	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN010621\
 Data File : VN065412.D
 Acq On : 6 Jan 2021 20:10
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Jan 07 01:31:19 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122220W.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 22 15:42:18 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.006	0.006	0.0	110	0.00
50 S	Toluene-d8	1.131	1.171	-3.5	119	0.00
51 T	4-Methyl-2-Pentanone	0.437	0.407	6.9	104	0.00
52 CM	Toluene	0.852	0.864	-1.4#	115	0.00
53 T	t-1,3-Dichloropropene	0.567	0.520	8.3	104	0.00
54 T	cis-1,3-Dichloropropene	0.603	0.568	5.8	108	0.00
55 T	1,1,2-Trichloroethane	0.337	0.348	-3.3	117	0.00
56 T	Ethyl methacrylate	0.522	0.504	3.4	109	0.00
57 T	1,3-Dichloropropane	0.557	0.574	-3.1	116	0.00
58 T	2-Chloroethyl Vinyl ether	0.227	0.210	7.5	107	0.00
59 T	2-Hexanone	0.321	0.294	8.4	101	0.00
60 T	Dibromochloromethane	0.402	0.390	3.0	110	0.00
61 T	1,2-Dibromoethane	0.362	0.364	-0.6	114	0.00
62 S	4-Bromofluorobenzene	0.424	0.429	-1.2	117	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	119	0.00
64 T	Tetrachloroethene	0.486	0.464	4.5	116	0.00
65 PM	Chlorobenzene	1.018	1.023	-0.5	118	0.00
66 T	1,1,1,2-Tetrachloroethane	0.396	0.388	2.0	114	0.00
67 C	Ethyl Benzene	1.834	1.777	3.1#	114	0.00
68 T	m/p-Xylenes	0.691	0.682	1.3	116	0.00
69 T	o-Xylene	0.671	0.659	1.8	115	0.00
70 T	Styrene	1.119	1.136	-1.5	117	0.00
71 P	Bromoform	0.341	0.320	6.2	108	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	121	0.00
73 T	Isopropylbenzene	3.808	3.506	7.9	115	0.00
74 T	N-amyl acetate	1.516	1.285	15.2	100	0.00
75 P	1,1,2,2-Tetrachloroethane	1.027	0.993	3.3	120	0.00
76 T	1,2,3-Trichloropropane	1.020	0.903	11.5	114	0.00
77 T	Bromobenzene	0.995	0.970	2.5	120	0.00
78 T	n-propylbenzene	4.299	3.908	9.1	112	0.00
79 T	2-Chlorotoluene	2.532	2.328	8.1	112	0.00
80 T	1,3,5-Trimethylbenzene	3.247	2.990	7.9	114	0.00
81 T	trans-1,4-Dichloro-2-butene	0.405	0.329	18.8	97	0.00
82 T	4-Chlorotoluene	2.607	2.423	7.1	114	0.00
83 T	tert-Butylbenzene	2.767	2.555	7.7	114	0.00
84 T	1,2,4-Trimethylbenzene	3.162	2.968	6.1	115	0.00
85 T	sec-Butylbenzene	3.671	3.333	9.2	112	0.00
86 T	p-Isopropyltoluene	3.365	3.092	8.1	113	0.00
87 T	1,3-Dichlorobenzene	1.664	1.653	0.7	120	0.00
88 T	1,4-Dichlorobenzene	1.671	1.640	1.9	119	0.00
89 T	n-Butylbenzene	2.876	2.536	11.8	106	0.00
90 T	Hexachloroethane	0.579	0.516	10.9	105	0.00
91 T	1,2-Dichlorobenzene	1.598	1.635	-2.3	122	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.212	0.187	11.8	103	0.00
93 T	1,2,4-Trichlorobenzene	0.806	0.899	-11.5	115	0.00
94 T	Hexachlorobutadiene	0.631	0.530	16.0	108	0.00
95 T	Naphthalene	2.105	2.287	-8.6	114	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN010621\
Data File : VN065412.D
Acq On : 6 Jan 2021 20:10
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: Jan 07 01:31:19 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122220W.M
Quant Title : SW846 8260
QLast Update : Tue Dec 22 15:42:18 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.764	0.854	-11.8	121	0.02

(#) = Out of Range SPCC's out = 0 CCC's out = 6