

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN010622\
 Data File : VN070511.D
 Acq On : 6 Jan 2022 11:42
 Operator : JC/MD
 Sample : VN0106WBS01
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :

Quant Time: Jan 08 01:35:58 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122721W.M
 Quant Title : SW846 8260
 QLast Update : Mon Dec 27 18:47:12 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.083	168	991021	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.965	114	1592756	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.741	117	1475210	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.672	152	632665	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.437	65	713597	50.081	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 100.160%	
35) Dibromofluoromethane	8.021	113	532559	55.262	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 110.520%	
50) Toluene-d8	10.440	98	2165681	55.011	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 110.020%	
62) 4-Bromofluorobenzene	12.728	95	781908	54.991	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 109.980%	

Target Compounds				Qvalue		
2) Dichlorodifluoromethane	2.073	85	212895	18.943	ug/l	99
3) Chloromethane	2.303	50	293238	19.181	ug/l	99
4) Vinyl Chloride	2.448	62	266666	19.158	ug/l	97
5) Bromomethane	2.864	94	150397	20.934	ug/l	97
6) Chloroethane	3.025	64	164663	19.644	ug/l	98
7) Trichlorofluoromethane	3.381	101	324720	19.382	ug/l	99
8) Diethyl Ether	3.824	74	149897	20.296	ug/l	75
9) 1,1,2-Trichlorotrifluo...	4.213	101	215685	21.049	ug/l	96
10) Methyl Iodide	4.422	142	271384	19.570	ug/l	98
11) Tert butyl alcohol	5.377	59	219611	92.564	ug/l	99
12) 1,1-Dichloroethene	4.183	96	200696	19.884	ug/l	92
13) Acrolein	4.038	56	122326	63.545	ug/l	97
14) Allyl chloride	4.843	41	470425	20.557	ug/l #	84
15) Acrylonitrile	5.551	53	790836	106.818	ug/l	99
16) Acetone	4.285	43	918558	99.306	ug/l	98
17) Carbon Disulfide	4.532	76	511372	18.561	ug/l	100
18) Methyl Acetate	4.856	43	549162	20.524	ug/l	92
19) Methyl tert-butyl Ether	5.615	73	780971	20.836	ug/l	94
20) Methylene Chloride	5.095	84	260518	20.357	ug/l	87
21) trans-1,2-Dichloroethene	5.597	96	218394	20.436	ug/l	87
22) Diisopropyl ether	6.495	45	1001879	22.478	ug/l	91
23) Vinyl Acetate	6.433	43	3662158	105.899	ug/l #	94
24) 1,1-Dichloroethane	6.385	63	482389	20.873	ug/l	98
25) 2-Butanone	7.335	43	1207665	103.692	ug/l	91
26) 2,2-Dichloropropane	7.324	77	349086	19.147	ug/l	98
27) cis-1,2-Dichloroethene	7.324	96	265469	20.819	ug/l	86
28) Bromochloromethane	7.656	49	235830	20.866	ug/l #	75
29) Tetrahydrofuran	7.683	42	740671	106.590	ug/l	87
30) Chloroform	7.817	83	457851	20.536	ug/l	100
31) Cyclohexane	8.094	56	477696	21.245	ug/l	89
32) 1,1,1-Trichloroethane	8.013	97	367118	19.529	ug/l	96
36) 1,1-Dichloropropene	8.220	75	337709	21.183	ug/l	96
37) Ethyl Acetate	7.410	43	446194	21.593	ug/l #	94
38) Carbon Tetrachloride	8.204	117	298776	19.627	ug/l	98
39) Methylcyclohexane	9.459	83	409241	21.204	ug/l	88
40) Benzene	8.458	78	1058701	21.661	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN010622\
 Data File : VN070511.D
 Acq On : 6 Jan 2022 11:42
 Operator : JC/MD
 Sample : VN0106WBS01
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :

Quant Time: Jan 08 01:35:58 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122721W.M
 Quant Title : SW846 8260
 QLast Update : Mon Dec 27 18:47:12 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.635	41	229086	22.164	ug/l	88
42) 1,2-Dichloroethane	8.528	62	371527	20.224	ug/l	99
43) Isopropyl Acetate	8.560	43	675473	20.733	ug/l #	96
44) Trichloroethene	9.215	130	241409	20.788	ug/l	91
45) 1,2-Dichloropropane	9.488	63	295721	22.472	ug/l	100
46) Dibromomethane	9.574	93	172284	21.054	ug/l	94
47) Bromodichloromethane	9.762	83	345056	21.237	ug/l	99
48) Methyl methacrylate	9.558	41	317026	20.400	ug/l #	81
49) 1,4-Dioxane	9.569	88	99043	435.389	ug/l #	80
51) 4-Methyl-2-Pentanone	10.328	43	2253782	110.092	ug/l	97
52) Toluene	10.502	92	647469	21.886	ug/l	98
53) t-1,3-Dichloropropene	10.717	75	360238	20.480	ug/l	99
54) cis-1,3-Dichloropropene	10.188	75	408311	21.459	ug/l	97
55) 1,1,2-Trichloroethane	10.896	97	257768	22.022	ug/l	97
56) Ethyl methacrylate	10.759	69	410758	21.238	ug/l	87
57) 1,3-Dichloropropane	11.041	76	453970	21.928	ug/l	100
58) 2-Chloroethyl Vinyl ether	10.038	63	474140	107.382	ug/l	90
59) 2-Hexanone	11.084	43	1630526	109.218	ug/l	96
60) Dibromochloromethane	11.234	129	242538	21.237	ug/l	100
61) 1,2-Dibromoethane	11.344	107	251476	21.159	ug/l	100
64) Tetrachloroethene	10.974	164	221128	20.700	ug/l	95
65) Chlorobenzene	11.768	112	672269	21.478	ug/l	99
66) 1,1,1,2-Tetrachloroethane	11.840	131	231899	20.753	ug/l	99
67) Ethyl Benzene	11.843	91	1251169	21.816	ug/l	96
68) m/p-Xylenes	11.950	106	931416	44.617	ug/l	94
69) o-Xylene	12.280	106	460336	22.170	ug/l	94
70) Styrene	12.291	104	739228	22.494	ug/l	97
71) Bromoform	12.455	173	175309	21.139	ug/l #	99
73) Isopropylbenzene	12.578	105	1189975	20.254	ug/l	99
74) N-amyl acetate	12.387	43	474499	19.208	ug/l	96
75) 1,1,2,2-Tetrachloroethane	12.825	83	408082	20.154	ug/l	98
76) 1,2,3-Trichloropropane	12.878	75	470166	26.464	ug/l	76
77) Bromobenzene	12.859	156	273778	20.297	ug/l	79
78) n-propylbenzene	12.919	91	1370860	21.059	ug/l	96
79) 2-Chlorotoluene	13.004	91	845234	20.237	ug/l	93
80) 1,3,5-Trimethylbenzene	13.058	105	980676	20.829	ug/l	98
81) trans-1,4-Dichloro-2-b...	12.623	75	99855	19.283	ug/l	96
82) 4-Chlorotoluene	13.104	91	804114	20.553	ug/l	92
83) tert-Butylbenzene	13.321	119	849623	20.494	ug/l	95
84) 1,2,4-Trimethylbenzene	13.366	105	949694	20.948	ug/l	100
85) sec-Butylbenzene	13.498	105	1169177	20.773	ug/l	97
86) p-Isopropyltoluene	13.613	119	934553	21.137	ug/l	97
87) 1,3-Dichlorobenzene	13.613	146	468153	20.733	ug/l	97
88) 1,4-Dichlorobenzene	13.694	146	451672	20.278	ug/l	97
89) n-Butylbenzene	13.940	91	727272	19.721	ug/l	97
90) Hexachloroethane	14.209	117	150346	19.574	ug/l	91
91) 1,2-Dichlorobenzene	13.986	146	456540	20.624	ug/l	97
92) 1,2-Dibromo-3-Chloropr...	14.600	75	61658	18.116	ug/l	82
93) 1,2,4-Trichlorobenzene	15.263	180	200045	19.179	ug/l	98
94) Hexachlorobutadiene	15.367	225	134229	22.143	ug/l	97
95) Naphthalene	15.504	128	457135	17.085	ug/l	99
96) 1,2,3-Trichlorobenzene	15.700	180	187726	19.332	ug/l	98


Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN010622\
Data File : VN070511.D
Acq On : 6 Jan 2022 11:42
Operator : JC/MD
Sample : VN0106WBS01
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :

Quant Time: Jan 08 01:35:58 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122721W.M
Quant Title : SW846 8260
QLast Update : Mon Dec 27 18:47:12 2021
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed



#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 8.083 min Scan# 21

Delta R.T. -0.003 min

Lab File: VN070511.D

Acq: 6 Jan 2022 11:42

Instrument :

MSVOA_N





ClientSampleId :

Tgt Ion:168 Resp: 991021

Ion Ratio Lower Upper

168 100

99 57.0 41.3 61.9



#2

Dichlorodifluoromethane

Concen: 18.943 ug/l

RT: 2.073 min Scan# 31

Delta R.T. -0.000 min

Lab File: VN070511.D




Acq: 6 Jan 2022 11:42


Tgt Ion: 85 Resp: 212895

Ion Ratio Lower Upper

85 100

87 31.8 16.3 48.9









#3
Chloromethane
Concen: 19.181 ug/l
RT: 2.303 min Scan# 117
Delta R.T. 0.000 min
Lab File: VN070511.D
Acq: 6 Jan 2022 11:42




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 50 Resp: 293238
Ion Ratio Lower Upper
50 100
52 32.3 25.4 38.0



#4
Vinyl Chloride
Concen: 19.158 ug/l
RT: 2.448 min Scan# 171
Delta R.T. 0.000 min
Lab File: VN070511.D
Acq: 6 Jan 2022 11:42

Tgt Ion: 62 Resp: 266666
Ion Ratio Lower Upper
62 100
64 30.6 25.8 38.6









#5
Bromomethane
Concen: 20.934 ug/l
RT: 2.864 min Scan# 31
Delta R.T. 0.011 min
Lab File: VN070511.D
Acq: 6 Jan 2022 11:42




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 94 Resp: 150397
Ion Ratio Lower Upper
94 100
96 94.9 73.8 110.8



#6
Chloroethane
Concen: 19.644 ug/l
RT: 3.025 min Scan# 386
Delta R.T. 0.006 min
Lab File: VN070511.D
Acq: 6 Jan 2022 11:42

Tgt Ion: 64 Resp: 164663
Ion Ratio Lower Upper
64 100
66 31.5 26.3 39.5









#7
 Trichlorofluoromethane
 Concen: 19.382 ug/l
 RT: 3.381 min Scan# 517
 Delta R.T. 0.005 min
 Lab File: VN070511.D
 Acq: 6 Jan 2022 11:42




Instrument :
 MSVOA_N
 ClientSampleId :

Tgt Ion:101 Resp: 324720
 Ion Ratio Lower Upper
 101 100
 103 63.5 50.1 75.1



#8
 Diethyl Ether
 Concen: 20.296 ug/l
 RT: 3.824 min Scan# 684
 Delta R.T. -0.000 min
 Lab File: VN070511.D
 Acq: 6 Jan 2022 11:42


Tgt Ion: 74 Resp: 149897
 Ion Ratio Lower Upper
 74 100
 45 117.8 47.0 141.2








#9
1,1,2-Trichlorotrifluoroethane
Concen: 21.049 ug/l
RT: 4.213 min Scan# 81
Delta R.T. 0.003 min
Lab File: VN070511.D
Acq: 6 Jan 2022 11:42

Instrument :
MSVOA_N
ClientSampleId :






Tgt Ion:101 Resp: 215685
Ion Ratio Lower Upper
101 100
85 44.9 37.2 55.8
151 77.9 65.9 98.9



#10
Methyl Iodide
Concen: 19.570 ug/l
RT: 4.422 min Scan# 907
Delta R.T. 0.000 min
Lab File: VN070511.D
Acq: 6 Jan 2022 11:42

Tgt Ion:142 Resp: 271384
Ion Ratio Lower Upper
142 100
127 42.1 35.4 53.0
141 14.4 11.6 17.4









#11
Tert butyl alcohol
Concen: 92.564 ug/l
RT: 5.377 min Scan# 11
Delta R.T. -0.000 min
Lab File: VN070511.D
Acq: 6 Jan 2022 11:42




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 59 Resp: 219611
Ion Ratio Lower Upper
59 100
57 10.4 8.7 13.1



#12
1,1-Dichloroethene
Concen: 19.884 ug/l
RT: 4.183 min Scan# 818
Delta R.T. 0.002 min
Lab File: VN070511.D
Acq: 6 Jan 2022 11:42


Tgt Ion: 96 Resp: 200696
Ion Ratio Lower Upper
96 100
61 176.7 131.0 196.6
98 63.4 52.4 78.6








#13
Acrolein
Concen: 63.545 ug/l
RT: 4.038 min Scan# 764
Delta R.T. -0.003 min
Lab File: VN070511.D
Acq: 6 Jan 2022 11:42

Instrument :
MSVOA_N
ClientSampleId :







Tgt Ion: 56 Resp: 122326
Ion Ratio Lower Upper
56 100
55 73.0 56.6 85.0



#14
Allyl chloride
Concen: 20.557 ug/l
RT: 4.843 min Scan# 1064
Delta R.T. 0.003 min
Lab File: VN070511.D
Acq: 6 Jan 2022 11:42

Tgt Ion: 41 Resp: 470425
Ion Ratio Lower Upper
41 100
39 60.7 61.2 91.8#
76 26.8 26.0 39.0









#15
Acrylonitrile
Concen: 106.818 ug/l
RT: 5.551 min Scan# 11
Delta R.T. -0.003 min
Lab File: VN070511.D
Acq: 6 Jan 2022 11:42




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 53 Resp: 790836
Ion Ratio Lower Upper
53 100
52 82.5 66.2 99.2
51 35.6 29.6 44.4



#16
Acetone
Concen: 99.306 ug/l
RT: 4.285 min Scan# 856
Delta R.T. 0.000 min
Lab File: VN070511.D
Acq: 6 Jan 2022 11:42

Tgt Ion: 43 Resp: 918558
Ion Ratio Lower Upper
43 100
58 30.2 23.1 34.7









#17
Carbon Disulfide
Concen: 18.561 ug/l
RT: 4.532 min Scan# 947
Delta R.T. 0.003 min
Lab File: VN070511.D
Acq: 6 Jan 2022 11:42




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 76 Resp: 511372
Ion Ratio Lower Upper
76 100
78 9.3 7.5 11.3



#18
Methyl Acetate
Concen: 20.524 ug/l
RT: 4.856 min Scan# 1069
Delta R.T. 0.002 min
Lab File: VN070511.D
Acq: 6 Jan 2022 11:42

Tgt Ion: 43 Resp: 549162
Ion Ratio Lower Upper
43 100
74 20.1 19.4 29.2





#19

Methyl tert-butyl Ether

Concen: 20.836 ug/l

RT: 5.615 min Scan# 11

Delta R.T. -0.001 min

Lab File: VN070511.D

Acq: 6 Jan 2022 11:42

Instrument :

MSVOA_N





ClientSampleId :

Tgt Ion: 73 Resp: 780971

Ion Ratio Lower Upper

73 100

57 24.0 16.8 25.2



#20

Methylene Chloride

Concen: 20.357 ug/l

RT: 5.095 min Scan# 1158

Delta R.T. 0.000 min

Lab File: VN070511.D

Acq: 6 Jan 2022 11:42

Tgt Ion: 84 Resp: 260518




Ion Ratio Lower Upper


84 100

49 139.2 93.5 140.3

51 42.5 29.4 44.2


86 65.3 53.0 79.4








#21
trans-1,2-Dichloroethene
Concen: 20.436 ug/l
RT: 5.597 min Scan# 11
Delta R.T. -0.000 min
Lab File: VN070511.D
Acq: 6 Jan 2022 11:42

Instrument :
MSVOA_N
ClientSampleId :







Tgt Ion: 96 Resp: 218394
Ion Ratio Lower Upper
96 100
61 151.3 105.8 158.6
98 60.2 52.0 78.0



#22
Diisopropyl ether
Concen: 22.478 ug/l
RT: 6.495 min Scan# 1680
Delta R.T. -0.003 min
Lab File: VN070511.D
Acq: 6 Jan 2022 11:42


Tgt Ion: 45 Resp: 1001879
Ion Ratio Lower Upper
45 100
43 57.3 51.7 77.5
87 23.6 23.2 34.8
59 10.4 9.4 14.0








#23
 Vinyl Acetate
 Concen: 105.899 ug/l
 RT: 6.433 min Scan# 1657
 Delta R.T. -0.001 min
 Lab File: VN070511.D
 Acq: 6 Jan 2022 11:42

Instrument :
 MSVOA_N
 ClientSampleId :






Tgt Ion: 43 Resp: 3662158
 Ion Ratio Lower Upper
 43 100
 86 8.3 8.5 12.7#



#24
 1,1-Dichloroethane
 Concen: 20.873 ug/l
 RT: 6.385 min Scan# 1639
 Delta R.T. -0.003 min
 Lab File: VN070511.D
 Acq: 6 Jan 2022 11:42

Tgt Ion: 63 Resp: 482389
 Ion Ratio Lower Upper
 63 100
 98 6.0 3.3 9.9
 100 3.4 2.1 6.5









#25
2-Butanone
Concen: 103.692 ug/l
RT: 7.335 min Scan# 1993
Delta R.T. -0.000 min
Lab File: VN070511.D
Acq: 6 Jan 2022 11:42




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 43 Resp: 1207665
Ion Ratio Lower Upper
43 100
72 22.5 21.7 32.5



#26
2,2-Dichloropropane
Concen: 19.147 ug/l
RT: 7.324 min Scan# 1989
Delta R.T. -0.003 min
Lab File: VN070511.D
Acq: 6 Jan 2022 11:42


Tgt Ion: 77 Resp: 349086
Ion Ratio Lower Upper
77 100
97 22.3 11.6 34.8








#27
 cis-1,2-Dichloroethene
 Concen: 20.819 ug/l
 RT: 7.324 min Scan# 1989
 Delta R.T. -0.000 min
 Lab File: VN070511.D
 Acq: 6 Jan 2022 11:42

Instrument :
 MSVOA_N
 ClientSampleId :



Tgt Ion: 96 Resp: 265469




Ion	Ratio	Lower	Upper
96	100		
61	160.6	0.0	275.0
98	63.4	0.0	129.6




#28
 Bromochloromethane
 Concen: 20.866 ug/l
 RT: 7.656 min Scan# 2113
 Delta R.T. -0.001 min
 Lab File: VN070511.D
 Acq: 6 Jan 2022 11:42

Tgt Ion: 49 Resp: 235830

Ion	Ratio	Lower	Upper
49	100		
129	1.7	0.0	4.0
130	61.4	67.8	101.8





#29

Tetrahydrofuran

Concen: 106.590 ug/l

RT: 7.683 min Scan# 2123

Delta R.T. 0.000 min


Lab File: VN070511.D

Acq: 6 Jan 2022 11:42

Instrument :

MSVOA_N

ClientSampleId :






Tgt Ion: 42 Resp: 740671

Ion Ratio Lower Upper

42 100

72 36.5 36.3 54.5

71 34.7 34.7 52.1



#30

Chloroform

Concen: 20.536 ug/l

RT: 7.817 min Scan# 2173

Delta R.T. 0.000 min

Lab File: VN070511.D




Acq: 6 Jan 2022 11:42


Tgt Ion: 83 Resp: 457851

Ion Ratio Lower Upper

83 100


85 64.6 51.9 77.9








#31
Cyclohexane
Concen: 21.245 ug/l
RT: 8.094 min Scan# 21
Delta R.T. -0.000 min
Lab File: VN070511.D
Acq: 6 Jan 2022 11:42

Instrument :
MSVOA_N
ClientSampleId :







Tgt Ion: 56 Resp: 477696
Ion Ratio Lower Upper
56 100
69 33.1 27.1 40.7
84 77.5 73.4 110.0



#32
1,1,1-Trichloroethane
Concen: 19.529 ug/l
RT: 8.013 min Scan# 2246
Delta R.T. 0.000 min
Lab File: VN070511.D
Acq: 6 Jan 2022 11:42

Tgt Ion: 97 Resp: 367118
Ion Ratio Lower Upper
97 100
99 63.0 50.1 75.1
61 49.2 35.2 52.8









#33
 1,2-Dichloroethane-d4
 Concen: 50.081 ug/l
 RT: 8.437 min Scan# 2404
 Delta R.T. -0.000 min
 Lab File: VN070511.D
 Acq: 6 Jan 2022 11:42

Instrument :
 MSVOA_N
 ClientSampleId :

Tgt Ion: 65 Resp: 713597




Ion	Ratio	Lower	Upper
65	100		
67	53.7	0.0	103.0




#34
 1,4-Difluorobenzene
 Concen: 50.000 ug/l
 RT: 8.965 min Scan# 2601
 Delta R.T. -0.003 min
 Lab File: VN070511.D
 Acq: 6 Jan 2022 11:42

Tgt Ion: 114 Resp: 1592756


Ion	Ratio	Lower	Upper
114	100		
63	23.2	0.0	37.8
88	16.3	0.0	27.4








#35
Dibromofluoromethane
Concen: 55.262 ug/l
RT: 8.021 min Scan# 211
Delta R.T. 0.000 min
Lab File: VN070511.D
Acq: 6 Jan 2022 11:42


Instrument :
MSVOA_N
ClientSampleId :





Tgt Ion:113 Resp: 532559
Ion Ratio Lower Upper
113 100
111 102.9 82.2 123.2
192 20.0 16.9 25.3




#36
1,1-Dichloropropene
Concen: 21.183 ug/l
RT: 8.220 min Scan# 2323
Delta R.T. -0.000 min
Lab File: VN070511.D
Acq: 6 Jan 2022 11:42



Tgt Ion: 75 Resp: 337709
Ion Ratio Lower Upper
75 100
110 32.6 18.4 55.2
77 31.0 24.4 36.6









#37
Ethyl Acetate
Concen: 21.593 ug/l
RT: 7.410 min Scan# 20
Delta R.T. -0.002 min
Lab File: VN070511.D
Acq: 6 Jan 2022 11:42




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 43 Resp: 446194
Ion Ratio Lower Upper
43 100
61 11.4 11.5 17.3#
70 8.3 7.9 11.9



#38
Carbon Tetrachloride
Concen: 19.627 ug/l
RT: 8.204 min Scan# 2317
Delta R.T. -0.002 min
Lab File: VN070511.D
Acq: 6 Jan 2022 11:42

Tgt Ion: 117 Resp: 298776
Ion Ratio Lower Upper
117 100
119 97.2 76.1 114.1
121 30.6 24.2 36.2





#39

Methylcyclohexane

Concen: 21.204 ug/l

RT: 9.459 min Scan# 21

Delta R.T. -0.000 min

Lab File: VN070511.D

Acq: 6 Jan 2022 11:42

Instrument :

MSVOA_N

ClientSampleId :





Tgt Ion: 83 Resp: 409241

Ion Ratio Lower Upper

83 100

55 86.8 58.1 87.1

98 44.1 38.4 57.6



#40

Benzene

Concen: 21.661 ug/l

RT: 8.458 min Scan# 2412

Delta R.T. 0.000 min

Lab File: VN070511.D




Acq: 6 Jan 2022 11:42


Tgt Ion: 78 Resp: 1058701

Ion Ratio Lower Upper

78 100


77 23.5 18.6 27.8








#41
Methacrylonitrile
Concen: 22.164 ug/l
RT: 7.635 min Scan# 2106
Delta R.T. -0.003 min
Lab File: VN070511.D
Acq: 6 Jan 2022 11:42

Instrument :
MSVOA_N
ClientSampleId :







Tgt Ion:	41	Resp:	229086
Ion	Ratio	Lower	Upper
41	100		
39	56.9	48.1	72.1
67	70.1	68.3	102.5
52	29.7	29.6	44.4



#42
1,2-Dichloroethane
Concen: 20.224 ug/l
RT: 8.528 min Scan# 2438
Delta R.T. 0.000 min
Lab File: VN070511.D
Acq: 6 Jan 2022 11:42


Tgt Ion:	62	Resp:	371527
Ion	Ratio	Lower	Upper
62	100		
98	8.5	0.0	17.6








#43
Isopropyl Acetate
Concen: 20.733 ug/l
RT: 8.560 min Scan# 2450
Delta R.T. 0.000 min
Lab File: VN070511.D
Acq: 6 Jan 2022 11:42

Instrument :
MSVOA_N
ClientSampleId :






Tgt Ion: 43 Resp: 675473
Ion Ratio Lower Upper
43 100
61 24.4 20.4 30.6
87 10.8 11.2 16.8#



#44
Trichloroethene
Concen: 20.788 ug/l
RT: 9.215 min Scan# 2694
Delta R.T. -0.000 min
Lab File: VN070511.D
Acq: 6 Jan 2022 11:42

Tgt Ion:130 Resp: 241409
Ion Ratio Lower Upper
130 100
95 103.9 0.0 191.2





#45

1,2-Dichloropropane

Concen: 22.472 ug/l

RT: 9.488 min Scan# 21

Delta R.T. 0.000 min

Lab File: VN070511.D

Acq: 6 Jan 2022 11:42

Instrument :

MSVOA_N





ClientSampleId :

Tgt Ion: 63 Resp: 295721

Ion Ratio Lower Upper

63 100

65 30.6 24.7 37.1



#46

Dibromomethane

Concen: 21.054 ug/l

RT: 9.574 min Scan# 2828

Delta R.T. -0.003 min

Lab File: VN070511.D

Acq: 6 Jan 2022 11:42




Tgt Ion: 93 Resp: 172284


Ion Ratio Lower Upper

93 100

95 84.6 65.9 98.9

174 96.3 84.6 127.0









#47
Bromodichloromethane
Concen: 21.237 ug/l
RT: 9.762 min Scan# 21
Delta R.T. -0.000 min
Lab File: VN070511.D
Acq: 6 Jan 2022 11:42




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 83 Resp: 345056
Ion Ratio Lower Upper
83 100
85 63.7 50.9 76.3
127 7.7 7.1 10.7



#48
Methyl methacrylate
Concen: 20.400 ug/l
RT: 9.558 min Scan# 2822
Delta R.T. -0.000 min
Lab File: VN070511.D
Acq: 6 Jan 2022 11:42

Tgt Ion: 41 Resp: 317026
Ion Ratio Lower Upper
41 100
69 77.2 72.4 108.6
39 51.4 58.4 87.6#





#49

1,4-Dioxane

Concen: 435.389 ug/l

RT: 9.569 min Scan# 2826

Delta R.T. -0.000 min


Lab File: VN070511.D

Acq: 6 Jan 2022 11:42

Instrument :

MSVOA_N

ClientSampleId :






Tgt Ion: 88 Resp: 99043

Ion Ratio Lower Upper

88 100

43 33.9 23.4 35.2

58 82.9 50.6 75.8#



#50

Toluene-d8

Concen: 55.011 ug/l

RT: 10.440 min Scan# 3151

Delta R.T. 0.000 min

Lab File: VN070511.D




Acq: 6 Jan 2022 11:42

Tgt Ion: 98 Resp: 2165681

Ion Ratio Lower Upper

98 100

100 64.2 50.4 75.6









#51
4-Methyl-2-Pentanone
Concen: 110.092 ug/l
RT: 10.328 min Scan# 3110
Delta R.T. -0.002 min
Lab File: VN070511.D
Acq: 6 Jan 2022 11:42




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 43 Resp: 2253782
Ion Ratio Lower Upper
43 100
58 36.2 30.4 45.6



#52
Toluene
Concen: 21.886 ug/l
RT: 10.502 min Scan# 3174
Delta R.T. -0.003 min
Lab File: VN070511.D
Acq: 6 Jan 2022 11:42

Tgt Ion: 92 Resp: 647469
Ion Ratio Lower Upper
92 100
91 174.4 142.0 213.0





#53

t-1,3-Dichloropropene

Concen: 20.480 ug/l

RT: 10.717 min Scan# 31

Delta R.T. -0.000 min

Lab File: VN070511.D

Acq: 6 Jan 2022 11:42

Instrument :

MSVOA_N





ClientSampleId :

Tgt Ion: 75 Resp: 360238

Ion Ratio Lower Upper

75 100

77 31.5 25.6 38.4



#54

cis-1,3-Dichloropropene

Concen: 21.459 ug/l

RT: 10.188 min Scan# 3057

Delta R.T. 0.000 min

Lab File: VN070511.D

Acq: 6 Jan 2022 11:42




Tgt Ion: 75 Resp: 408311


Ion Ratio Lower Upper

75 100

77 31.0 26.4 39.6

39 54.7 42.5 63.7









#55
1,1,2-Trichloroethane
Concen: 22.022 ug/l
RT: 10.896 min Scan# 31
Delta R.T. 0.000 min
Lab File: VN070511.D
Acq: 6 Jan 2022 11:42




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 97 Resp: 257768
Ion Ratio Lower Upper
97 100
83 89.7 67.8 101.6
85 55.1 45.4 68.2
99 62.3 49.4 74.0



#56
Ethyl methacrylate
Concen: 21.238 ug/l
RT: 10.759 min Scan# 3270
Delta R.T. -0.001 min
Lab File: VN070511.D
Acq: 6 Jan 2022 11:42

Tgt Ion: 69 Resp: 410758
Ion Ratio Lower Upper
69 100
41 76.1 51.4 77.0
39 37.8 34.6 52.0





#57

1,3-Dichloropropane

Concen: 21.928 ug/l

RT: 11.041 min Scan# 31

Delta R.T. -0.003 min

Lab File: VN070511.D

Acq: 6 Jan 2022 11:42

Instrument :

MSVOA_N





ClientSampleId :

Tgt Ion: 76 Resp: 453970

Ion Ratio Lower Upper

76 100

78 31.9 25.7 38.5



#58

2-Chloroethyl Vinyl ether

Concen: 107.382 ug/l

RT: 10.038 min Scan# 3001

Delta R.T. -0.003 min

Lab File: VN070511.D




Acq: 6 Jan 2022 11:42


Tgt Ion: 63 Resp: 474140

Ion Ratio Lower Upper

63 100

106 24.5 23.9 35.9





#59

2-Hexanone

Concen: 109.218 ug/l

RT: 11.084 min Scan# 3391

Delta R.T. 0.000 min

Lab File: VN070511.D

Acq: 6 Jan 2022 11:42

Instrument :

MSVOA_N





ClientSampleId :

Tgt Ion: 43 Resp: 1630526

Ion Ratio Lower Upper

43 100

58 50.6 26.9 80.7



#60

Dibromochloromethane

Concen: 21.237 ug/l

RT: 11.234 min Scan# 3447

Delta R.T. -0.003 min

Lab File: VN070511.D




Acq: 6 Jan 2022 11:42

Tgt Ion:129 Resp: 242538

Ion Ratio Lower Upper

129 100

127 77.1 38.6 116.0









#61
1,2-Dibromoethane
Concen: 21.159 ug/l
RT: 11.344 min Scan# 3488
Delta R.T. 0.000 min
Lab File: VN070511.D
Acq: 6 Jan 2022 11:42




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion:107 Resp: 251476
Ion Ratio Lower Upper
107 100
109 94.4 75.7 113.5



#62
4-Bromofluorobenzene
Concen: 54.991 ug/l
RT: 12.728 min Scan# 4004
Delta R.T. -0.003 min
Lab File: VN070511.D
Acq: 6 Jan 2022 11:42


Tgt Ion: 95 Resp: 781908
Ion Ratio Lower Upper
95 100
174 75.7 0.0 187.2
176 73.7 0.0 181.4








#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.741 min Scan# 3636
Delta R.T. -0.003 min
Lab File: VN070511.D
Acq: 6 Jan 2022 11:42


Instrument :
MSVOA_N
ClientSampleId :





Tgt Ion:117 Resp: 1475210
Ion Ratio Lower Upper
117 100
82 58.9 42.4 63.6
119 32.2 25.4 38.2




#64
Tetrachloroethene
Concen: 20.700 ug/l
RT: 10.974 min Scan# 3350
Delta R.T. -0.003 min
Lab File: VN070511.D
Acq: 6 Jan 2022 11:42



Tgt Ion:164 Resp: 221128
Ion Ratio Lower Upper
164 100
166 129.2 101.9 152.9
129 97.6 72.0 108.0
131 93.5 70.9 106.3









#65
Chlorobenzene
Concen: 21.478 ug/l
RT: 11.768 min Scan# 3646
Delta R.T. -0.000 min
Lab File: VN070511.D
Acq: 6 Jan 2022 11:42




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion:112 Resp: 672269
Ion Ratio Lower Upper
112 100
114 32.1 25.4 38.0



#66
1,1,1,2-Tetrachloroethane
Concen: 20.753 ug/l
RT: 11.840 min Scan# 3673
Delta R.T. 0.000 min
Lab File: VN070511.D
Acq: 6 Jan 2022 11:42

Tgt Ion:131 Resp: 231899
Ion Ratio Lower Upper
131 100
133 94.4 47.9 143.7
119 65.5 32.1 96.5









#67
Ethyl Benzene
Concen: 21.816 ug/l
RT: 11.843 min Scan# 3674
Delta R.T. 0.000 min
Lab File: VN070511.D
Acq: 6 Jan 2022 11:42




Instrument :
MSVOA_N
ClientSampleId :

Tgt Ion: 91 Resp: 1251169
Ion Ratio Lower Upper
91 100
106 29.8 25.7 38.5



#68
m/p-Xylenes
Concen: 44.617 ug/l
RT: 11.950 min Scan# 3714
Delta R.T. -0.003 min
Lab File: VN070511.D
Acq: 6 Jan 2022 11:42

Tgt Ion:106 Resp: 931416
Ion Ratio Lower Upper
106 100
91 210.0 160.6 240.8









#69
o-Xylene
Concen: 22.170 ug/l
RT: 12.280 min Scan# 3837
Delta R.T. 0.002 min
Lab File: VN070511.D
Acq: 6 Jan 2022 11:42




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion:106 Resp: 460336
Ion Ratio Lower Upper
106 100
91 221.2 105.8 317.3



#70
Styrene
Concen: 22.494 ug/l
RT: 12.291 min Scan# 3841
Delta R.T. -0.003 min
Lab File: VN070511.D
Acq: 6 Jan 2022 11:42

Tgt Ion:104 Resp: 739228
Ion Ratio Lower Upper
104 100
78 52.9 38.7 58.1
103 55.8 45.0 67.6









#71
Bromoform
Concen: 21.139 ug/l
RT: 12.455 min Scan# 3902
Delta R.T. -0.002 min
Lab File: VN070511.D
Acq: 6 Jan 2022 11:42




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion:173 Resp: 175309
Ion Ratio Lower Upper
173 100
175 47.7 24.2 72.6
254 0.0 0.0 0.0



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.672 min Scan# 4356
Delta R.T. 0.000 min
Lab File: VN070511.D
Acq: 6 Jan 2022 11:42


Tgt Ion:152 Resp: 632665
Ion Ratio Lower Upper
152 100
115 59.9 27.5 82.5
150 163.0 0.0 344.4








#73
Isopropylbenzene
Concen: 20.254 ug/l
RT: 12.578 min Scan# 3948
Delta R.T. -0.000 min
Lab File: VN070511.D
Acq: 6 Jan 2022 11:42

Instrument :
MSVOA_N
ClientSampleId :







Tgt Ion:105 Resp: 1189975
Ion Ratio Lower Upper
105 100
120 26.0 13.4 40.1



#74
N-ethyl acetate
Concen: 19.208 ug/l
RT: 12.387 min Scan# 3877
Delta R.T. 0.002 min
Lab File: VN070511.D
Acq: 6 Jan 2022 11:42

Tgt Ion: 43 Resp: 474499
Ion Ratio Lower Upper
43 100
70 37.1 33.4 50.0
55 23.8 19.3 28.9
61 21.6 18.5 27.7





#75

1,1,2,2-Tetrachloroethane

Concen: 20.154 ug/l

RT: 12.825 min Scan# 4040

Delta R.T. -0.000 min

Lab File: VN070511.D

Acq: 6 Jan 2022 11:42

Instrument :

MSVOA_N

ClientSampleId :





Tgt Ion: 83 Resp: 408082

Ion Ratio Lower Upper

83 100

131 9.5 5.5 16.5

85 63.1 32.4 97.2



#76

1,2,3-Trichloropropane

Concen: 26.464 ug/l

RT: 12.878 min Scan# 4060

Delta R.T. 0.000 min

Lab File: VN070511.D




Acq: 6 Jan 2022 11:42


Tgt Ion: 75 Resp: 470166

Ion Ratio Lower Upper

75 100

77 132.2 82.6 247.8









#77
Bromobenzene
Concen: 20.297 ug/l
RT: 12.859 min Scan# 4053
Delta R.T. -0.001 min
Lab File: VN070511.D
Acq: 6 Jan 2022 11:42




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion:156 Resp: 273778
Ion Ratio Lower Upper
156 100
77 227.1 90.6 271.8
158 97.3 48.9 146.7



#78
n-propylbenzene
Concen: 21.059 ug/l
RT: 12.919 min Scan# 4075
Delta R.T. -0.001 min
Lab File: VN070511.D
Acq: 6 Jan 2022 11:42

Tgt Ion: 91 Resp: 1370860
Ion Ratio Lower Upper
91 100
120 21.8 11.9 35.9









#79
2-Chlorotoluene
Concen: 20.237 ug/l
RT: 13.004 min Scan# 4108
Delta R.T. -0.003 min
Lab File: VN070511.D
Acq: 6 Jan 2022 11:42




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 91 Resp: 845234
Ion Ratio Lower Upper
91 100
126 32.5 18.3 54.8



#80
1,3,5-Trimethylbenzene
Concen: 20.829 ug/l
RT: 13.058 min Scan# 4127
Delta R.T. -0.000 min
Lab File: VN070511.D
Acq: 6 Jan 2022 11:42

Tgt Ion:105 Resp: 980676
Ion Ratio Lower Upper
105 100
120 47.5 24.6 73.6









#81
trans-1,4-Dichloro-2-butene
Concen: 19.283 ug/l
RT: 12.623 min Scan# 3965
Delta R.T. -0.001 min
Lab File: VN070511.D
Acq: 6 Jan 2022 11:42




Instrument :
MSVOA_N
ClientSampleId :

Tgt Ion: 75 Resp: 99855
Ion Ratio Lower Upper
75 100
53 116.4 90.5 135.7
89 44.3 38.2 57.4



#82
4-Chlorotoluene
Concen: 20.553 ug/l
RT: 13.104 min Scan# 4144
Delta R.T. -0.000 min
Lab File: VN070511.D
Acq: 6 Jan 2022 11:42

Tgt Ion: 91 Resp: 804114
Ion Ratio Lower Upper
91 100
126 31.9 18.2 54.6









#83
 tert-Butylbenzene
 Concen: 20.494 ug/l
 RT: 13.321 min Scan# 41
 Delta R.T. -0.000 min
 Lab File: VN070511.D
 Acq: 6 Jan 2022 11:42

Instrument :
 MSVOA_N
 ClientSampleId :

Tgt Ion:119 Resp: 849623




Ion	Ratio	Lower	Upper
119	100		
91	67.5	31.5	94.5
134	23.4	12.4	37.4




#84
 1,2,4-Trimethylbenzene
 Concen: 20.948 ug/l
 RT: 13.366 min Scan# 4242
 Delta R.T. 0.000 min
 Lab File: VN070511.D
 Acq: 6 Jan 2022 11:42

Tgt Ion:105 Resp: 949694

Ion	Ratio	Lower	Upper
105	100		
120	45.3	22.8	68.4





#85

sec-Butylbenzene

Concen: 20.773 ug/l

RT: 13.498 min Scan# 4291

Delta R.T. -0.003 min

Lab File: VN070511.D

Acq: 6 Jan 2022 11:42

Instrument :

MSVOA_N





ClientSampleId :

Tgt Ion:105 Resp: 1169177

Ion Ratio Lower Upper

105 100

134 19.4 10.4 31.1



#86

p-Isopropyltoluene

Concen: 21.137 ug/l

RT: 13.613 min Scan# 4334

Delta R.T. 0.000 min

Lab File: VN070511.D

Acq: 6 Jan 2022 11:42




Tgt Ion:119 Resp: 934553


Ion Ratio Lower Upper

119 100

134 25.5 13.6 40.7

91 23.9 11.4 34.1









#87
1,3-Dichlorobenzene
Concen: 20.733 ug/l
RT: 13.613 min Scan# 4334
Delta R.T. 0.000 min
Lab File: VN070511.D
Acq: 6 Jan 2022 11:42




Instrument :
MSVOA_N
ClientSampleId :

Tgt Ion:146 Resp: 468153
Ion Ratio Lower Upper
146 100
111 41.7 18.6 55.6
148 64.0 31.6 95.0



#88
1,4-Dichlorobenzene
Concen: 20.278 ug/l
RT: 13.694 min Scan# 4364
Delta R.T. -0.000 min
Lab File: VN070511.D
Acq: 6 Jan 2022 11:42

Tgt Ion:146 Resp: 451672
Ion Ratio Lower Upper
146 100
111 41.6 18.3 54.8
148 63.8 31.8 95.4









#89
n-Butylbenzene
Concen: 19.721 ug/l
RT: 13.940 min Scan# 4456
Delta R.T. 0.000 min
Lab File: VN070511.D
Acq: 6 Jan 2022 11:42




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 91 Resp: 727272
Ion Ratio Lower Upper
91 100
92 51.8 25.4 76.0
134 24.7 13.6 40.8



#90
Hexachloroethane
Concen: 19.574 ug/l
RT: 14.209 min Scan# 4556
Delta R.T. -0.000 min
Lab File: VN070511.D
Acq: 6 Jan 2022 11:42


Tgt Ion: 117 Resp: 150346
Ion Ratio Lower Upper
117 100
201 79.6 43.9 131.6








#91
1,2-Dichlorobenzene
Concen: 20.624 ug/l
RT: 13.986 min Scan# 4473
Delta R.T. -0.000 min
Lab File: VN070511.D
Acq: 6 Jan 2022 11:42

Instrument :
MSVOA_N
ClientSampleId :







Tgt Ion:146 Resp: 456540
Ion Ratio Lower Upper
146 100
111 42.4 19.1 57.3
148 63.8 32.1 96.3



#92
1,2-Dibromo-3-Chloropropane
Concen: 18.116 ug/l
RT: 14.600 min Scan# 4702
Delta R.T. 0.000 min
Lab File: VN070511.D
Acq: 6 Jan 2022 11:42

Tgt Ion: 75 Resp: 61658
Ion Ratio Lower Upper
75 100
155 77.8 46.9 140.7
157 99.0 60.6 181.7





#93

1,2,4-Trichlorobenzene

Concen: 19.179 ug/l

RT: 15.263 min Scan# 4949

Delta R.T. -0.000 min

Lab File: VN070511.D

Acq: 6 Jan 2022 11:42

Instrument :

MSVOA_N

ClientSampleId :





Tgt Ion:180 Resp: 200045

Ion Ratio Lower Upper

180 100

182 93.7 48.3 144.9

145 29.3 14.5 43.5



#94

Hexachlorobutadiene

Concen: 22.143 ug/l

RT: 15.367 min Scan# 4988

Delta R.T. 0.000 min

Lab File: VN070511.D

Acq: 6 Jan 2022 11:42




Tgt Ion:225 Resp: 134229


Ion Ratio Lower Upper

225 100

223 62.1 31.9 95.7

227 61.8 32.1 96.3









#95
Naphthalene
Concen: 17.085 ug/l
RT: 15.504 min Scan# 5039
Delta R.T. -0.000 min
Lab File: VN070511.D
Acq: 6 Jan 2022 11:42

Instrument :
MSVOA_N
ClientSampleId :

Tgt Ion:128 Resp: 457135
Ion Ratio Lower Upper
128 100
127 12.6 10.5 15.7
129 10.7 8.7 13.1



#96
1,2,3-Trichlorobenzene
Concen: 19.332 ug/l
RT: 15.700 min Scan# 5112
Delta R.T. -0.000 min
Lab File: VN070511.D
Acq: 6 Jan 2022 11:42

Tgt Ion:180 Resp: 187726
Ion Ratio Lower Upper
180 100
182 96.2 47.4 142.2
145 32.2 15.0 45.1

