

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN010622\
 Data File : VN070509.D
 Acq On : 6 Jan 2022 10:48
 Operator : JC/MD
 Sample : VN0106MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0106MBL01

Quant Time: Jan 08 01:35:17 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122721W.M
 Quant Title : SW846 8260
 QLast Update : Mon Dec 27 18:47:12 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.085	168	963569	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.965	114	1862797	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.744	117	2229671	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.675	152	898614	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.437	65	692256	49.967	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	99.940%
35) Dibromofluoromethane	8.021	113	541251	48.022	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	96.040%
50) Toluene-d8	10.440	98	2486025	53.994	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	107.980%
62) 4-Bromofluorobenzene	12.731	95	1078743	64.869	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	129.740%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN010622\
Data File : VN070509.D
Acq On : 6 Jan 2022 10:48
Operator : JC/MD
Sample : VN0106MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0106MBL01

Quant Time: Jan 08 01:35:17 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122721W.M
Quant Title : SW846 8260
QLast Update : Mon Dec 27 18:47:12 2021
Response via : Initial Calibration

