

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN010622\
 Data File : VN070510.D
 Acq On : 6 Jan 2022 11:13
 Operator : JC/MD
 Sample : VN0106WBL01
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0106WBL01

Quant Time: Jan 08 01:35:39 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122721W.M
 Quant Title : SW846 8260
 QLast Update : Mon Dec 27 18:47:12 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.083	168	951594	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.965	114	1600642	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.744	117	1374617	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.672	152	450478	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.437	65	731801	53.486	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	106.980%
35) Dibromofluoromethane	8.021	113	524781	54.187	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	108.380%
50) Toluene-d8	10.440	98	2065647	52.212	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	104.420%
62) 4-Bromofluorobenzene	12.731	95	705309	49.359	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	98.720%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN010622\
Data File : VN070510.D
Acq On : 6 Jan 2022 11:13
Operator : JC/MD
Sample : VN0106WBL01
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0106WBL01

Quant Time: Jan 08 01:35:39 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122721W.M
Quant Title : SW846 8260
QLast Update : Mon Dec 27 18:47:12 2021
Response via : Initial Calibration

