

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN010622\
 Data File : VN070518.D
 Acq On : 6 Jan 2022 14:46
 Operator : JC/MD
 Sample : N1012-08
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 31240

Quant Time: Jan 08 01:38:56 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122721W.M
 Quant Title : SW846 8260
 QLast Update : Mon Dec 27 18:47:12 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	783906	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.965	114	1314643	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.744	117	1162090	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.675	152	397042	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.437	65	611538	54.257	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 108.520%	
35) Dibromofluoromethane	8.021	113	431974	54.307	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 108.620%	
50) Toluene-d8	10.440	98	1717808	52.866	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 105.740%	
62) 4-Bromofluorobenzene	12.731	95	622636	53.053	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 106.100%	
Target Compounds						
						Qvalue
16) Acetone	4.288	43	226052	30.895	ug/l	97
25) 2-Butanone	7.343	43	9326	1.012	ug/l	91
39) Methylcyclohexane	9.454	83	5206	0.327	ug/l	98
40) Benzene	8.461	78	17157	0.425	ug/l	97
68) m/p-Xylenes	11.953	106	8620	0.524	ug/l	100
84) 1,2,4-Trimethylbenzene	13.369	105	11396	0.401	ug/l	100
95) Naphthalene	15.507	128	36252	3.729	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN010622\
Data File : VN070518.D
Acq On : 6 Jan 2022 14:46
Operator : JC/MD
Sample : N1012-08
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
31240

Quant Time: Jan 08 01:38:56 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122721W.M
Quant Title : SW846 8260
QLast Update : Mon Dec 27 18:47:12 2021
Response via : Initial Calibration

