

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN010622\
 Data File : VN070520.D
 Acq On : 6 Jan 2022 15:36
 Operator : JC/MD
 Sample : N1012-13
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 31245

Quant Time: Jan 08 01:39:36 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122721W.M
 Quant Title : SW846 8260
 QLast Update : Mon Dec 27 18:47:12 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.083	168	883024	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.965	114	1494365	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.744	117	1348655	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.672	152	461609	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.434	65	657161	51.760	ug/l	0.00
Spiked Amount	50.000	Range 61 - 141	Recovery	= 103.520%		
35) Dibromofluoromethane	8.021	113	460803	50.964	ug/l	0.00
Spiked Amount	50.000	Range 69 - 133	Recovery	= 101.920%		
50) Toluene-d8	10.440	98	1867556	50.562	ug/l	0.00
Spiked Amount	50.000	Range 65 - 126	Recovery	= 101.120%		
62) 4-Bromofluorobenzene	12.731	95	685654	51.396	ug/l	0.00
Spiked Amount	50.000	Range 58 - 135	Recovery	= 102.800%		
Target Compounds						
						Qvalue
11) Tert butyl alcohol	5.380	59	8652	4.093	ug/l	# 83
16) Acetone	4.288	43	308106	37.383	ug/l	97
25) 2-Butanone	7.343	43	17999	1.734	ug/l	96
39) Methylcyclohexane	9.456	83	6331	0.350	ug/l	93
40) Benzene	8.458	78	56135	1.224	ug/l	100
52) Toluene	10.502	92	48899	1.762	ug/l	95
68) m/p-Xylenes	11.953	106	6340	0.332	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN010622\
Data File : VN070520.D
Acq On : 6 Jan 2022 15:36
Operator : JC/MD
Sample : N1012-13
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
31245

Quant Time: Jan 08 01:39:36 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122721W.M
Quant Title : SW846 8260
QLast Update : Mon Dec 27 18:47:12 2021
Response via : Initial Calibration

