

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN010622\
 Data File : VN070522.D
 Acq On : 6 Jan 2022 16:27
 Operator : JC/MD
 Sample : N1012-07DL 40X
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 31239DL

Quant Time: Jan 08 01:40:15 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122721W.M
 Quant Title : SW846 8260
 QLast Update : Mon Dec 27 18:47:12 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	924209	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.968	114	1572465	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.744	117	1365646	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.675	152	454916	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.437	65	611012	45.981	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	91.960%		
35) Dibromofluoromethane	8.018	113	431638	45.368	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	90.740%		
50) Toluene-d8	10.440	98	1719857	44.250	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	88.500%		
62) 4-Bromofluorobenzene	12.731	95	606885	43.232	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	86.460%		
Target Compounds					Qvalue	
16) Acetone	4.288	43	414210	48.018	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN010622\
Data File : VN070522.D
Acq On : 6 Jan 2022 16:27
Operator : JC/MD
Sample : N1012-07DL 40X
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
31239DL

Quant Time: Jan 08 01:40:15 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122721W.M
Quant Title : SW846 8260
QLast Update : Mon Dec 27 18:47:12 2021
Response via : Initial Calibration

