

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN010622\
 Data File : VN070527.D
 Acq On : 6 Jan 2022 18:34
 Operator : JC/MD
 Sample : PB141860TB
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB141860TB

Quant Time: Jan 08 01:41:46 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122721W.M
 Quant Title : SW846 8260
 QLast Update : Mon Dec 27 18:47:12 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	869905	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.968	114	1497656	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.744	117	1352794	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.675	152	459692	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.437	65	676320	54.073	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 108.140%	
35) Dibromofluoromethane	8.021	113	475421	52.466	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 104.940%	
50) Toluene-d8	10.440	98	1956460	52.852	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 105.700%	
62) 4-Bromofluorobenzene	12.731	95	698801	52.267	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 104.540%	
Target Compounds						
						Qvalue
16) Acetone	4.288	43	113333	13.958	ug/l	100
20) Methylene Chloride	5.087	84	62803	5.591	ug/l	90
43) Isopropyl Acetate	8.563	43	142463	4.650	ug/l #	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN010622\
Data File : VN070527.D
Acq On : 6 Jan 2022 18:34
Operator : JC/MD
Sample : PB141860TB
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB141860TB

Quant Time: Jan 08 01:41:46 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122721W.M
Quant Title : SW846 8260
QLast Update : Mon Dec 27 18:47:12 2021
Response via : Initial Calibration

