

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN010622\
 Data File : VN070528.D
 Acq On : 6 Jan 2022 18:59
 Operator : JC/MD
 Sample : PB141877TB
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB141877TB

Quant Time: Jan 08 01:42:06 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122721W.M
 Quant Title : SW846 8260
 QLast Update : Mon Dec 27 18:47:12 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	878080	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.965	114	1545391	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.744	117	1368086	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.675	152	462100	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.437	65	685287	54.280	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	108.560%
35) Dibromofluoromethane	8.021	113	480689	51.408	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	102.820%
50) Toluene-d8	10.440	98	1983518	51.928	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	103.860%
62) 4-Bromofluorobenzene	12.728	95	702426	50.915	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	101.820%
Target Compounds						
					Qvalue	
16) Acetone	4.293	43	114883	14.018	ug/l	98
18) Methyl Acetate	4.865	43	27452	1.158	ug/l #	78
20) Methylene Chloride	5.093	84	64146	5.657	ug/l	89
43) Isopropyl Acetate	8.560	43	160146	5.066	ug/l #	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN010622\
Data File : VN070528.D
Acq On : 6 Jan 2022 18:59
Operator : JC/MD
Sample : PB141877TB
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB141877TB

Quant Time: Jan 08 01:42:06 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122721W.M
Quant Title : SW846 8260
QLast Update : Mon Dec 27 18:47:12 2021
Response via : Initial Calibration

