

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN010622\
 Data File : VN070534.D
 Acq On : 6 Jan 2022 21:31
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Jan 08 01:44:19 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122721W.M
 Quant Title : SW846 8260
 QLast Update : Mon Dec 27 18:47:12 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	79	0.00
2 T	Dichlorodifluoromethane	0.567	0.506	10.8	67	0.00
3 P	Chloromethane	0.771	0.735	4.7	77	0.00
4 C	Vinyl Chloride	0.702	0.666	5.1#	75	0.00
5 T	Bromomethane	0.362	0.359	0.8	73	0.00
6 T	Chloroethane	0.423	0.405	4.3	76	0.00
7 T	Trichlorofluoromethane	0.845	0.798	5.6	74	0.00
8 T	Diethyl Ether	0.373	0.386	-3.5	82	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.517	0.519	-0.4	78	0.00
10 T	Methyl Iodide	0.700	0.716	-2.3	76	0.00
11 T	Tert butyl alcohol	0.120	0.110	8.3	72	0.00
12 CM	1,1-Dichloroethene	0.509	0.498	2.2#	76	0.00
13 T	Acrolein	0.097	0.076	21.6	63	0.00
14 T	Allyl chloride	1.155	1.194	-3.4	81	0.00
15 T	Acrylonitrile	0.374	0.410	-9.6	85	0.00
16 T	Acetone	0.467	0.361	22.7	58	0.00
17 T	Carbon Disulfide	1.390	1.260	9.4	71	0.00
18 T	Methyl Acetate	1.350	1.445	-7.0	85	0.00
19 T	Methyl tert-butyl Ether	1.891	1.992	-5.3	81	0.00
20 T	Methylene Chloride	0.646	0.645	0.2	82	0.00
21 T	trans-1,2-Dichloroethene	0.539	0.537	0.4	77	0.00
22 T	Diisopropyl ether	2.249	2.471	-9.9	86	0.00
23 T	Vinyl Acetate	1.745	1.916	-9.8	83	0.00
24 P	1,1-Dichloroethane	1.166	1.200	-2.9	81	0.00
25 T	2-Butanone	0.588	0.572	2.7	74	0.00
26 T	2,2-Dichloropropane	0.920	0.832	9.6	70	0.00
27 T	cis-1,2-Dichloroethene	0.643	0.666	-3.6	80	0.00
28 T	Bromochloromethane	0.570	0.647	-13.5	89	0.00
29 T	Tetrahydrofuran	0.351	0.385	-9.7	83	0.00
30 C	Chloroform	1.125	1.131	-0.5#	79	0.00
31 T	Cyclohexane	1.134	1.139	-0.4	79	0.00
32 T	1,1,1-Trichloroethane	0.948	0.924	2.5	75	0.00
33 S	1,2-Dichloroethane-d4	0.719	0.701	2.5	77	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	80	0.00
35 S	Dibromofluoromethane	0.303	0.311	-2.6	81	0.00
36 T	1,1-Dichloropropene	0.500	0.494	1.2	78	0.00
37 T	Ethyl Acetate	0.649	0.678	-4.5	82	0.00
38 T	Carbon Tetrachloride	0.478	0.443	7.3	72	0.00
39 T	Methylcyclohexane	0.606	0.607	-0.2	77	0.00
40 TM	Benzene	1.534	1.582	-3.1	81	0.00
41 T	Methacrylonitrile	0.324	0.366	-13.0	86	0.00
42 TM	1,2-Dichloroethane	0.577	0.551	4.5	77	0.00
43 T	Isopropyl Acetate	1.023	1.102	-7.7	85	0.00
44 TM	Trichloroethene	0.365	0.357	2.2	78	0.00
45 C	1,2-Dichloropropane	0.413	0.447	-8.2#	85	0.00
46 T	Dibromomethane	0.257	0.258	-0.4	79	0.00
47 T	Bromodichloromethane	0.510	0.520	-2.0	78	0.00
48 T	Methyl methacrylate	0.488	0.540	-10.7	81	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN010622\
 Data File : VN070534.D
 Acq On : 6 Jan 2022 21:31
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Jan 08 01:44:19 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122721W.M
 Quant Title : SW846 8260
 QLast Update : Mon Dec 27 18:47:12 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.007	0.007	0.0	79	0.00
50 S	Toluene-d8	1.236	1.272	-2.9	82	0.00
51 T	4-Methyl-2-Pentanone	0.643	0.702	-9.2	83	0.00
52 CM	Toluene	0.929	0.968	-4.2#	81	0.00
53 T	t-1,3-Dichloropropene	0.552	0.572	-3.6	78	0.00
54 T	cis-1,3-Dichloropropene	0.597	0.637	-6.7	80	0.00
55 T	1,1,2-Trichloroethane	0.367	0.385	-4.9	82	0.00
56 T	Ethyl methacrylate	0.607	0.674	-11.0	84	0.00
57 T	1,3-Dichloropropane	0.650	0.694	-6.8	84	0.00
58 T	2-Chloroethyl Vinyl ether	0.127	0.162	-27.6#	93	0.00
59 T	2-Hexanone	0.469	0.500	-6.6	79	0.00
60 T	Dibromochloromethane	0.359	0.373	-3.9	78	0.00
61 T	1,2-Dibromoethane	0.373	0.388	-4.0	81	0.00
62 S	4-Bromofluorobenzene	0.446	0.474	-6.3	83	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	80	0.00
64 T	Tetrachloroethene	0.362	0.325	10.2	73	0.00
65 PM	Chlorobenzene	1.061	1.082	-2.0	81	0.00
66 T	1,1,1,2-Tetrachloroethane	0.379	0.381	-0.5	79	0.00
67 C	Ethyl Benzene	1.944	2.023	-4.1#	80	0.00
68 T	m/p-Xylenes	0.708	0.748	-5.6	81	0.00
69 T	o-Xylene	0.704	0.753	-7.0	81	0.00
70 T	Styrene	1.114	1.235	-10.9	82	0.00
71 P	Bromoform	0.281	0.299	-6.4	79	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	83	0.00
73 T	Isopropylbenzene	4.643	4.435	4.5	81	0.00
74 T	N-amyl acetate	1.952	2.041	-4.6	87	0.00
75 P	1,1,2,2-Tetrachloroethane	1.559	1.523	2.3	85	0.00
76 T	1,2,3-Trichloropropane	1.501	1.360	9.4	85	0.00
77 T	Bromobenzene	1.066	1.016	4.7	81	0.00
78 T	n-propylbenzene	5.145	5.075	1.4	81	0.00
79 T	2-Chlorotoluene	3.301	3.123	5.4	81	0.00
80 T	1,3,5-Trimethylbenzene	3.721	3.616	2.8	80	0.00
81 T	trans-1,4-Dichloro-2-butene	0.409	0.425	-3.9	81	0.00
82 T	4-Chlorotoluene	3.092	2.983	3.5	81	0.00
83 T	tert-Butylbenzene	3.276	3.207	2.1	82	0.00
84 T	1,2,4-Trimethylbenzene	3.583	3.551	0.9	80	0.00
85 T	sec-Butylbenzene	4.448	4.453	-0.1	82	0.00
86 T	p-Isopropyltoluene	3.494	3.542	-1.4	81	0.00
87 T	1,3-Dichlorobenzene	1.785	1.745	2.2	83	0.00
88 T	1,4-Dichlorobenzene	1.760	1.695	3.7	83	0.00
89 T	n-Butylbenzene	2.915	2.943	-1.0	82	0.00
90 T	Hexachloroethane	0.607	0.593	2.3	78	0.00
91 T	1,2-Dichlorobenzene	1.749	1.707	2.4	84	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.269	0.258	4.1	80	0.00
93 T	1,2,4-Trichlorobenzene	0.824	0.833	-1.1	86	0.00
94 T	Hexachlorobutadiene	0.479	0.491	-2.5	87	0.00
95 T	Naphthalene	2.240	2.317	-3.4	87	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN010622\
Data File : VN070534.D
Acq On : 6 Jan 2022 21:31
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: Jan 08 01:44:19 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122721W.M
Quant Title : SW846 8260
QLast Update : Mon Dec 27 18:47:12 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.804	0.815	-1.4	87	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6