

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN010622\
 Data File : VN070509.D
 Acq On : 6 Jan 2022 10:48
 Operator : JC/MD
 Sample : VN0106MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0106MBL01

Quant Time: Jan 08 01:35:17 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122721W.M
 Quant Title : SW846 8260
 QLast Update : Mon Dec 27 18:47:12 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.085	168	963569	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.965	114	1862797	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.744	117	2229671	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.675	152	898614	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.437	65	692256	49.967	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	99.940%
35) Dibromofluoromethane	8.021	113	541251	48.022	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	96.040%
50) Toluene-d8	10.440	98	2486025	53.994	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	107.980%
62) 4-Bromofluorobenzene	12.731	95	1078743	64.869	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	129.740%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN010622\
Data File : VN070509.D
Acq On : 6 Jan 2022 10:48
Operator : JC/MD
Sample : VN0106MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0106MBL01

Quant Time: Jan 08 01:35:17 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122721W.M
Quant Title : SW846 8260
QLast Update : Mon Dec 27 18:47:12 2021
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.085 min Scan# 21
Delta R.T. -0.001 min
Lab File: VN070509.D
Acq: 6 Jan 2022 10:48




Instrument :
MSVOA_N
ClientSampleId :
VN0106MBL01


Tgt Ion:168 Resp: 963569
Ion Ratio Lower Upper
168 100
99 56.9 41.3 61.9



#33
1,2-Dichloroethane-d4
Concen: 49.967 ug/l
RT: 8.437 min Scan# 2404
Delta R.T. -0.000 min
Lab File: VN070509.D
Acq: 6 Jan 2022 10:48

Tgt Ion: 65 Resp: 692256
Ion Ratio Lower Upper
65 100
67 53.8 0.0 103.0





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.965 min Scan# 2

Delta R.T. -0.003 min

Lab File: VN070509.D

Acq: 6 Jan 2022 10:48

Instrument :

MSVOA_N

ClientSampleId :

VN0106MBL01





Tgt Ion:114 Resp: 1862797

Ion Ratio Lower Upper

114 100

63 23.3 0.0 37.8

88 16.7 0.0 27.4



#35

Dibromofluoromethane

Concen: 48.022 ug/l

RT: 8.021 min Scan# 2249

Delta R.T. 0.000 min

Lab File: VN070509.D

Acq: 6 Jan 2022 10:48




Tgt Ion:113 Resp: 541251


Ion Ratio Lower Upper

113 100

111 103.3 82.2 123.2

192 19.8 16.9 25.3









#50
Toluene-d8
Concen: 53.994 ug/l
RT: 10.440 min Scan# 3151
Delta R.T. 0.000 min
Lab File: VN070509.D
Acq: 6 Jan 2022 10:48




Instrument :
MSVOA_N
ClientSampleId :
VN0106MBL01


Tgt Ion: 98 Resp: 2486025
Ion Ratio Lower Upper
98 100
100 64.3 50.4 75.6



#62
4-Bromofluorobenzene
Concen: 64.869 ug/l
RT: 12.731 min Scan# 4005
Delta R.T. -0.000 min
Lab File: VN070509.D
Acq: 6 Jan 2022 10:48

Tgt Ion: 95 Resp: 1078743
Ion Ratio Lower Upper
95 100
174 76.3 0.0 187.2
176 73.2 0.0 181.4





#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 11.744 min Scan# 3

Delta R.T. -0.000 min

Lab File: VN070509.D

Acq: 6 Jan 2022 10:48

Instrument :

MSVOA_N

ClientSampleId :

VN0106MBL01





Tgt Ion:117 Resp: 2229671

Ion Ratio Lower Upper

117 100

82 59.7 42.4 63.6

119 32.0 25.4 38.2



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.675 min Scan# 4357

Delta R.T. 0.003 min

Lab File: VN070509.D

Acq: 6 Jan 2022 10:48

Tgt Ion:152 Resp: 898614

Ion Ratio Lower Upper

152 100

115 60.5 27.5 82.5

150 157.1 0.0 344.4

