

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN010622\
 Data File : VN070510.D
 Acq On : 6 Jan 2022 11:13
 Operator : JC/MD
 Sample : VN0106WBL01
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0106WBL01

Quant Time: Jan 08 01:35:39 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122721W.M
 Quant Title : SW846 8260
 QLast Update : Mon Dec 27 18:47:12 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.083	168	951594	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.965	114	1600642	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.744	117	1374617	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.672	152	450478	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.437	65	731801	53.486	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	106.980%
35) Dibromofluoromethane	8.021	113	524781	54.187	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	108.380%
50) Toluene-d8	10.440	98	2065647	52.212	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	104.420%
62) 4-Bromofluorobenzene	12.731	95	705309	49.359	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	98.720%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN010622\
Data File : VN070510.D
Acq On : 6 Jan 2022 11:13
Operator : JC/MD
Sample : VN0106WBL01
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0106WBL01

Quant Time: Jan 08 01:35:39 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122721W.M
Quant Title : SW846 8260
QLast Update : Mon Dec 27 18:47:12 2021
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.083 min Scan# 21
Delta R.T. -0.003 min
Lab File: VN070510.D
Acq: 6 Jan 2022 11:13




Instrument :
MSVOA_N
ClientSampleId :
VN0106WBL01


Tgt Ion:168 Resp: 951594
Ion Ratio Lower Upper
168 100
99 57.6 41.3 61.9



#33
1,2-Dichloroethane-d4
Concen: 53.486 ug/l
RT: 8.437 min Scan# 2404
Delta R.T. -0.000 min
Lab File: VN070510.D
Acq: 6 Jan 2022 11:13

Tgt Ion: 65 Resp: 731801
Ion Ratio Lower Upper
65 100
67 53.0 0.0 103.0





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.965 min Scan# 2601

Delta R.T. -0.003 min

Lab File: VN070510.D

Acq: 6 Jan 2022 11:13

Instrument :

MSVOA_N

ClientSampleId :

VN0106WBL01





Tgt Ion:114 Resp: 1600642

Ion Ratio Lower Upper

114 100

63 23.5 0.0 37.8

88 16.5 0.0 27.4



#35

Dibromofluoromethane

Concen: 54.187 ug/l

RT: 8.021 min Scan# 2249

Delta R.T. 0.000 min

Lab File: VN070510.D

Acq: 6 Jan 2022 11:13




Tgt Ion:113 Resp: 524781


Ion Ratio Lower Upper

113 100

111 103.0 82.2 123.2

192 19.3 16.9 25.3









#50
Toluene-d8
Concen: 52.212 ug/l
RT: 10.440 min Scan# 3151
Delta R.T. 0.000 min
Lab File: VN070510.D
Acq: 6 Jan 2022 11:13




Instrument :
MSVOA_N
ClientSampleId :
VN0106WBL01


Tgt Ion: 98 Resp: 2065647
Ion Ratio Lower Upper
98 100
100 64.6 50.4 75.6



#62
4-Bromofluorobenzene
Concen: 49.359 ug/l
RT: 12.731 min Scan# 4005
Delta R.T. -0.000 min
Lab File: VN070510.D
Acq: 6 Jan 2022 11:13

Tgt Ion: 95 Resp: 705309
Ion Ratio Lower Upper
95 100
174 76.0 0.0 187.2
176 73.9 0.0 181.4





#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 11.744 min Scan# 3

Delta R.T. -0.000 min

Lab File: VN070510.D


Acq: 6 Jan 2022 11:13

Instrument :

MSVOA_N

ClientSampleId :

VN0106WBL01






Tgt Ion:117 Resp: 1374617

Ion Ratio Lower Upper

117 100

82 58.4 42.4 63.6

119 32.2 25.4 38.2



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.672 min Scan# 4356

Delta R.T. 0.000 min

Lab File: VN070510.D

Acq: 6 Jan 2022 11:13

Tgt Ion:152 Resp: 450478

Ion Ratio Lower Upper

152 100

115 59.3 27.5 82.5

150 156.7 0.0 344.4

