

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN010623\
 Data File : VN076122.D
 Acq On : 06 Jan 2023 10:25
 Operator : JC\MD
 Sample : VN0106MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0106MBL01

Quant Time: Jan 06 22:53:02 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121422W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 14 13:03:41 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.235	168	274316	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	432302	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	366831	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	147453	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	147798	49.525	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	99.060%
35) Dibromofluoromethane	8.177	113	133604	49.905	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.800%
50) Toluene-d8	10.571	98	493333	49.286	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	98.580%
62) 4-Bromofluorobenzene	12.853	95	146240	44.110	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	88.220%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN010623\
Data File : VN076122.D
Acq On : 06 Jan 2023 10:25
Operator : JC\MD
Sample : VN0106MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0106MBL01

Quant Time: Jan 06 22:53:02 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121422W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 14 13:03:41 2022
Response via : Initial Calibration

