

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN010623\
 Data File : VN076127.D
 Acq On : 06 Jan 2023 12:36
 Operator : JC\MD
 Sample : 01049-03
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 EB-01-010523

Quant Time: Jan 06 22:54:30 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121422W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 14 13:03:41 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.229	168	328534	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	528374	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	461250	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	188699	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	174367	48.786	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	97.580%
35) Dibromofluoromethane	8.176	113	158012	48.290	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.580%
50) Toluene-d8	10.570	98	608040	49.700	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	99.400%
62) 4-Bromofluorobenzene	12.853	95	187925	46.377	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	92.760%
Target Compounds						
					Qvalue	
16) Acetone	4.441	43	15477	12.562	ug/l #	77
20) Methylene Chloride	5.288	84	4860	0.845	ug/l #	87
25) 2-Butanone	7.500	43	3055	1.771	ug/l #	68

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN010623\
Data File : VN076127.D
Acq On : 06 Jan 2023 12:36
Operator : JC\MD
Sample : 01049-03
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
EB-01-010523

Quant Time: Jan 06 22:54:30 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121422W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 14 13:03:41 2022
Response via : Initial Calibration

