

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN010623\
 Data File : VN076128.D
 Acq On : 06 Jan 2023 13:00
 Operator : JC\MD
 Sample : 01049-01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 GWOW-BR-04-340-361-010523

Quant Time: Jan 06 22:54:42 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121422W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 14 13:03:41 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	310429	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	499918	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	440246	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	177945	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	166328	49.251	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	98.500%		
35) Dibromofluoromethane	8.171	113	151373	48.894	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	97.780%		
50) Toluene-d8	10.571	98	577110	49.857	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	99.720%		
62) 4-Bromofluorobenzene	12.853	95	179718	46.876	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	93.760%		
Target Compounds						
25) 2-Butanone	7.494	43	2236	1.372	ug/l	99
52) Toluene	10.635	92	1739358	210.475	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN010623\
Data File : VN076128.D
Acq On : 06 Jan 2023 13:00
Operator : JC\MD
Sample : 01049-01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
GWOW-BR-04-340-361-010523

Quant Time: Jan 06 22:54:42 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121422W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 14 13:03:41 2022
Response via : Initial Calibration

