

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN010623\
 Data File : VN076129.D
 Acq On : 06 Jan 2023 13:24
 Operator : JC\MD
 Sample : 01049-04
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 GWOW-BR-04-366-387-010523

Quant Time: Jan 06 22:54:54 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121422W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 14 13:03:41 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	316799	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	509809	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	445152	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	179584	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	170540	49.482	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	98.960%
35) Dibromofluoromethane	8.177	113	154455	48.922	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.840%
50) Toluene-d8	10.571	98	584424	49.509	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	99.020%
62) 4-Bromofluorobenzene	12.853	95	177248	45.335	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	90.680%
Target Compounds						
52) Toluene	10.635	92	346797	41.151	ug/l	Qvalue 100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN010623\
Data File : VN076129.D
Acq On : 06 Jan 2023 13:24
Operator : JC\MD
Sample : 01049-04
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
GWOW-BR-04-366-387-010523

Quant Time: Jan 06 22:54:54 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121422W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 14 13:03:41 2022
Response via : Initial Calibration

