

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN010623\
 Data File : VN076131.D
 Acq On : 06 Jan 2023 14:12
 Operator : JC\MD
 Sample : 01029-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-205

Quant Time: Jan 06 22:55:21 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121422W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 14 13:03:41 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.235	168	332853	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	542811	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	487477	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	201229	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	178729	49.357	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	98.720%
35) Dibromofluoromethane	8.171	113	160771	47.826	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.660%
50) Toluene-d8	10.570	98	634849	50.511	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	101.020%
62) 4-Bromofluorobenzene	12.853	95	200556	48.178	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	96.360%
Target Compounds						
					Qvalue	
16) Acetone	4.453	43	39387	31.553	ug/l	97
20) Methylene Chloride	5.289	84	12688	3.200	ug/l	95
37) Ethyl Acetate	7.571	43	29718	8.285	ug/l	99
43) Isopropyl Acetate	8.694	43	207092	35.001	ug/l #	79

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN010623\
Data File : VN076131.D
Acq On : 06 Jan 2023 14:12
Operator : JC\MD
Sample : 01029-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-205

Quant Time: Jan 06 22:55:21 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121422W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 14 13:03:41 2022
Response via : Initial Calibration

