

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN010623\
 Data File : VN076132.D
 Acq On : 06 Jan 2023 14:36
 Operator : JC\MD
 Sample : 01029-05
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-204

Quant Time: Jan 06 22:55:39 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121422W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 14 13:03:41 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.235	168	317339	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	520876	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	475877	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	193356	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	172480	49.960	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	99.920%
35) Dibromofluoromethane	8.177	113	155600	48.237	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.480%
50) Toluene-d8	10.571	98	612100	50.752	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	101.500%
62) 4-Bromofluorobenzene	12.853	95	192869	48.283	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	96.560%
Target Compounds						
					Qvalue	
16) Acetone	4.447	43	38516	32.364	ug/l	98
20) Methylene Chloride	5.294	84	12638	3.373	ug/l	84
37) Ethyl Acetate	7.577	43	28905	8.398	ug/l	99
43) Isopropyl Acetate	8.694	43	242608	42.730	ug/l #	74

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN010623\
Data File : VN076132.D
Acq On : 06 Jan 2023 14:36
Operator : JC\MD
Sample : 01029-05
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-204

Quant Time: Jan 06 22:55:39 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121422W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 14 13:03:41 2022
Response via : Initial Calibration

