

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN010623\
 Data File : VN076135.D
 Acq On : 06 Jan 2023 15:48
 Operator : JC\MD
 Sample : 01043-05
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-202

Quant Time: Jan 06 22:56:25 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121422W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 14 13:03:41 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	324254	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	535356	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	481518	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	193923	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	175868	49.855	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	99.720%
35) Dibromofluoromethane	8.171	113	157991	47.654	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.300%
50) Toluene-d8	10.571	98	627351	50.610	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	101.220%
62) 4-Bromofluorobenzene	12.853	95	195608	47.644	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	95.280%
Target Compounds						
					Qvalue	
16) Acetone	4.442	43	40448	33.262	ug/l	88
20) Methylene Chloride	5.289	84	10451	2.606	ug/l	87
37) Ethyl Acetate	7.577	43	29740	8.407	ug/l	99
43) Isopropyl Acetate	8.694	43	398061	68.213	ug/l #	66

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN010623\
Data File : VN076135.D
Acq On : 06 Jan 2023 15:48
Operator : JC\MD
Sample : 01043-05
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-202

Quant Time: Jan 06 22:56:25 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121422W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 14 13:03:41 2022
Response via : Initial Calibration

