

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN010623\
 Data File : VN076138.D
 Acq On : 06 Jan 2023 17:00
 Operator : JC\MD
 Sample : 01055-03
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MH-F

Quant Time: Jan 06 22:57:06 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121422W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 14 13:03:41 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.236	168	325966	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	532196	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	467820	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	190053	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.589	65	176825	49.863	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	99.720%
35) Dibromofluoromethane	8.171	113	164055	49.777	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.560%
50) Toluene-d8	10.571	98	614313	49.852	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	99.700%
62) 4-Bromofluorobenzene	12.853	95	187972	46.056	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	92.120%
Target Compounds						
64) Tetrachloroethene	11.106	164	68137	20.698	ug/l	Qvalue 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN010623\
Data File : VN076138.D
Acq On : 06 Jan 2023 17:00
Operator : JC\MD
Sample : 01055-03
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MH-F

Quant Time: Jan 06 22:57:06 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121422W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 14 13:03:41 2022
Response via : Initial Calibration

