

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN010625\
 Data File : VN085381.D
 Acq On : 06 Jan 2025 11:34
 Operator : JC\MD
 Sample : VN0106MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0106MBL01

Quant Time: Jan 07 00:20:02 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122724W.M
 Quant Title : SW846 8260
 QLast Update : Sat Dec 28 01:36:16 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.218	168	131686	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	250029	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	221529	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	86097	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	118523	63.104	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 126.200%#		
35) Dibromofluoromethane	8.165	113	91361	58.572	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 117.140%		
50) Toluene-d8	10.565	98	307975	53.806	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 107.620%		
62) 4-Bromofluorobenzene	12.847	95	88667	46.644	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	= 93.280%		

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN010625\
Data File : VN085381.D
Acq On : 06 Jan 2025 11:34
Operator : JC\MD
Sample : VN0106MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0106MBL01

Quant Time: Jan 07 00:20:02 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122724W.M
Quant Title : SW846 8260
QLast Update : Sat Dec 28 01:36:16 2024
Response via : Initial Calibration

