

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN010625\
 Data File : VN085387.D
 Acq On : 06 Jan 2025 14:07
 Operator : JC\MD
 Sample : PB165935ZHE#01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB165935ZHE#01

Quant Time: Jan 07 00:23:46 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122724W.M
 Quant Title : SW846 8260
 QLast Update : Sat Dec 28 01:36:16 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.218	168	161521	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	321169	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	294898	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	109588	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.571	65	136050	59.056	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 118.120%			
35) Dibromofluoromethane	8.165	113	103640	51.726	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 103.460%			
50) Toluene-d8	10.565	98	393423	53.509	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 107.020%			
62) 4-Bromofluorobenzene	12.847	95	118302	48.449	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 96.900%			
Target Compounds						
16) Acetone	4.436	43	23007	30.371	ug/l	91
20) Methylene Chloride	5.277	84	3748	1.839	ug/l #	93
30) Chloroform	7.965	83	114968	30.136	ug/l	95
43) Isopropyl Acetate	8.682	43	133118	27.726	ug/l #	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN010625\
Data File : VN085387.D
Acq On : 06 Jan 2025 14:07
Operator : JC\MD
Sample : PB165935ZHE#01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB165935ZHE#01

Quant Time: Jan 07 00:23:46 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122724W.M
Quant Title : SW846 8260
QLast Update : Sat Dec 28 01:36:16 2024
Response via : Initial Calibration

