

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN010721\
 Data File : VN065420.D
 Acq On : 7 Jan 2021 12:22
 Operator : JC/MD
 Sample : M1037-04
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TB-01-010621

Quant Time: Jan 08 01:15:53 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122220W.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 22 15:42:18 2020
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.630	168	236009	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.553	114	390562	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.379	117	410008	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.318	152	153397	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.991	65	131436	49.33	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	98.66%
35) Dibromofluoromethane	7.553	113	120559	51.16	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	102.32%
50) Toluene-d8	10.058	98	487029	55.15	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	110.30%
62) 4-Bromofluorobenzene	12.376	95	177643	50.77	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	101.54%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN010721\
Data File : VN065420.D
Acq On : 7 Jan 2021 12:22
Operator : JC/MD
Sample : M1037-04
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TB-01-010621

Quant Time: Jan 08 01:15:53 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122220W.M
Quant Title : SW846 8260
QLast Update : Tue Dec 22 15:42:18 2020
Response via : Initial Calibration

