

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN010721\
 Data File : VN065421.D
 Acq On : 7 Jan 2021 12:47
 Operator : JC/MD
 Sample : M1037-03
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 EB-01-010621

Quant Time: Jan 08 02:25:53 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122220W.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 22 15:42:18 2020
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.631	168	262232	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.553	114	439380	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.380	117	467630	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.318	152	179108	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.991	65	148122	50.03	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 100.06%	
35) Dibromofluoromethane	7.553	113	135121	50.97	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 101.94%	
50) Toluene-d8	10.058	98	549809	55.34	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 110.68%	
62) 4-Bromofluorobenzene	12.376	95	205141	51.98	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 103.96%	
Target Compounds						Qvalue
20) Methylene Chloride	4.515	84	3164	1.10	ug/l	# 83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN010721\
Data File : VN065421.D
Acq On : 7 Jan 2021 12:47
Operator : JC/MD
Sample : M1037-03
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
EB-01-010621

Quant Time: Jan 08 02:25:53 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122220W.M
Quant Title : SW846 8260
QLast Update : Tue Dec 22 15:42:18 2020
Response via : Initial Calibration

