

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN010721\
 Data File : VN065423.D
 Acq On : 7 Jan 2021 13:36
 Operator : JC/MD
 Sample : M1037-02
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 GWBR2-63-68-010621-FD

Manual Integrations
 APPROVED

MMDadoda
 1/8/2021 11:02:57 AM

Quant Time: Jan 08 01:16:44 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122220W.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 22 15:42:18 2020
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.631	168	255727	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.553	114	431357	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.380	117	452722	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.318	152	176166	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.991	65	145737	50.48	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery = 100.96%			
35) Dibromofluoromethane	7.553	113	132201	50.80	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery = 101.60%			
50) Toluene-d8	10.061	98	534816	54.83	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery = 109.66%			
62) 4-Bromofluorobenzene	12.376	95	198753	51.37	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery = 102.74%			
Target Compounds						Qvalue
4) Vinyl Chloride	2.181	62	45372	16.19	ug/l	99
9) 1,1,2-Trichlorotrifluo...	3.721	101	3969	1.69	ug/l	97
12) 1,1-Dichloroethene	3.714	96	1067m	0.42	ug/l	
21) trans-1,2-Dichloroethene	5.000	96	1217	0.49	ug/l #	81
24) 1,1-Dichloroethane	5.807	63	35238	7.79	ug/l	98
27) cis-1,2-Dichloroethene	6.788	96	373610	130.18	ug/l	89
30) Chloroform	7.335	83	37004	7.95	ug/l	96
32) 1,1,1-Trichloroethane	7.531	97	9090	2.16	ug/l	97
40) Benzene	8.004	78	5323	0.46	ug/l	98
44) Trichloroethene	8.804	130	3347	0.99	ug/l	98
47) Bromodichloromethane	9.373	83	4691	1.08	ug/l	94
52) Toluene	10.122	92	4285	0.58	ug/l	100
64) Tetrachloroethene	10.602	164	31715	7.21	ug/l	95
65) Chlorobenzene	11.405	112	1912	0.21	ug/l	96
67) Ethyl Benzene	11.483	91	4121	0.25	ug/l	93
68) m/p-Xylenes	11.592	106	3397	0.54	ug/l	93
69) o-Xylene	11.923	106	1917	0.32	ug/l	91
84) 1,2,4-Trimethylbenzene	13.016	105	3470	0.31	ug/l	99
95) Naphthalene	15.109	128	6638	4.05	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN010721\
Data File : VN065423.D
Acq On : 7 Jan 2021 13:36
Operator : JC/MD
Sample : M1037-02
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
GWBR2-63-68-010621-FD

Manual Integrations
APPROVED

MMDadoda
1/8/2021 11:02:57 AM

Quant Time: Jan 08 01:16:44 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122220W.M
Quant Title : SW846 8260
QLast Update : Tue Dec 22 15:42:18 2020
Response via : Initial Calibration

