

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN010721\
 Data File : VN065425.D
 Acq On : 7 Jan 2021 14:26
 Operator : JC/MD
 Sample : M1037-03RE
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 EB-01-010621RE

Quant Time: Jan 08 01:17:35 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122220W.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 22 15:42:18 2020
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.630	168	259191	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.553	114	431772	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.379	117	456858	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.318	152	178337	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.990	65	146229	49.97	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	99.94%
35) Dibromofluoromethane	7.553	113	133765	51.35	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	102.70%
50) Toluene-d8	10.061	98	544270	55.75	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	111.50%
62) 4-Bromofluorobenzene	12.376	95	201717	52.01	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	104.02%
Target Compounds						
20) Methylene Chloride	4.508	84	3537	1.24	ug/l	Qvalue 93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN010721\
Data File : VN065425.D
Acq On : 7 Jan 2021 14:26
Operator : JC/MD
Sample : M1037-03RE
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
EB-01-010621RE

Quant Time: Jan 08 01:17:35 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122220W.M
Quant Title : SW846 8260
QLast Update : Tue Dec 22 15:42:18 2020
Response via : Initial Calibration

