

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN010721\
 Data File : VN065434.D
 Acq On : 7 Jan 2021 18:09
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 IBLK

Quant Time: Jan 08 01:20:36 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122220W.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 22 15:42:18 2020
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.630	168	315935	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.553	114	514540	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.379	117	534447	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.318	152	233374	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.994	65	174371	48.89	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery = 97.78%			
35) Dibromofluoromethane	7.553	113	157704	50.80	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery = 101.60%			
50) Toluene-d8	10.061	98	626425	53.84	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery = 107.68%			
62) 4-Bromofluorobenzene	12.376	95	252511	54.38	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery = 108.76%			
Target Compounds					Qvalue	
40) Benzene	8.003	78	4448	0.32	ug/l #	80
52) Toluene	10.129	92	4355	0.50	ug/l	97
68) m/p-Xylenes	11.592	106	3271	0.44	ug/l	99
69) o-Xylene	11.923	106	1839	0.26	ug/l	96
70) Styrene	11.939	104	3511	0.29	ug/l	98
95) Naphthalene	15.106	128	68535	8.66	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN010721\
Data File : VN065434.D
Acq On : 7 Jan 2021 18:09
Operator : JC/MD
Sample : IBLK
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
IBLK

Quant Time: Jan 08 01:20:36 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122220W.M
Quant Title : SW846 8260
QLast Update : Tue Dec 22 15:42:18 2020
Response via : Initial Calibration

