

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN010721\
 Data File : VN065415.D
 Acq On : 7 Jan 2021 9:56
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Jan 08 01:13:44 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122220W.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 22 15:42:18 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	122	0.00
2 T	Dichlorodifluoromethane	0.463	0.434	6.3	108	0.00
3 P	Chloromethane	0.527	0.531	-0.8	118	0.00
4 C	Vinyl Chloride	0.548	0.579	-5.7#	122	0.00
5 T	Bromomethane	0.389	0.404	-3.9	124	0.00
6 T	Chloroethane	0.354	0.380	-7.3	125	0.00
7 T	Trichlorofluoromethane	0.820	0.787	4.0	113	0.00
8 T	Diethyl Ether	0.323	0.316	2.2	120	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.460	0.465	-1.1	122	0.00
10 T	Methyl Iodide	0.694	0.706	-1.7	117	0.00
11 T	Tert butyl alcohol	0.092	0.072	21.7	92	0.00
12 CM	1,1-Dichloroethene	0.491	0.471	4.1#	118	0.00
13 T	Acrolein	0.065	0.060	7.7	115	0.00
14 T	Allyl chloride	0.810	0.716	11.6	105	0.00
15 T	Acrylonitrile	0.223	0.217	2.7	113	0.00
16 T	Acetone	0.260	0.241	7.3	121	0.00
17 T	Carbon Disulfide	1.257	1.116	11.2	103	0.00
18 T	Methyl Acetate	0.507	0.495	2.4	115	0.00
19 T	Methyl tert-butyl Ether	1.560	1.469	5.8	113	0.00
20 T	Methylene Chloride	0.551	0.519	5.8	122	0.00
21 T	trans-1,2-Dichloroethene	0.482	0.480	0.4	118	0.00
22 T	Diisopropyl ether	1.611	1.544	4.2	116	0.00
23 T	Vinyl Acetate	1.358	1.238	8.8	106	0.00
24 P	1,1-Dichloroethane	0.884	0.887	-0.3	118	0.00
25 T	2-Butanone	0.317	0.296	6.6	109	0.00
26 T	2,2-Dichloropropane	0.844	0.767	9.1	111	0.00
27 T	cis-1,2-Dichloroethene	0.561	0.586	-4.5	124	0.00
28 T	Bromochloromethane	0.379	0.352	7.1	113	0.00
29 T	Tetrahydrofuran	0.205	0.178	13.2	105	0.00
30 C	Chloroform	0.910	0.922	-1.3#	121	0.00
31 T	Cyclohexane	0.833	0.734	11.9	110	0.00
32 T	1,1,1-Trichloroethane	0.824	0.779	5.5	112	0.00
33 S	1,2-Dichloroethane-d4	0.564	0.482	14.5	103	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	117	0.00
35 S	Dibromofluoromethane	0.302	0.300	0.7	114	0.00
36 T	1,1-Dichloropropene	0.455	0.466	-2.4	118	0.00
37 T	Ethyl Acetate	0.451	0.414	8.2	104	0.00
38 T	Carbon Tetrachloride	0.477	0.463	2.9	112	0.00
39 T	Methylcyclohexane	0.546	0.541	0.9	111	0.00
40 TM	Benzene	1.352	1.448	-7.1	124	0.00
41 T	Methacrylonitrile	0.234	0.215	8.1	111	0.00
42 TM	1,2-Dichloroethane	0.500	0.483	3.4	113	0.00
43 T	Isopropyl Acetate	0.777	0.685	11.8	100	0.00
44 TM	Trichloroethene	0.390	0.422	-8.2	124	0.00
45 C	1,2-Dichloropropane	0.352	0.382	-8.5#	124	0.00
46 T	Dibromomethane	0.239	0.246	-2.9	119	0.00
47 T	Bromodichloromethane	0.503	0.507	-0.8	115	0.00
48 T	Methyl methacrylate	0.377	0.331	12.2	101	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN010721\
 Data File : VN065415.D
 Acq On : 7 Jan 2021 9:56
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Jan 08 01:13:44 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122220W.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 22 15:42:18 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.006	0.006	0.0	115	0.00
50 S	Toluene-d8	1.131	1.127	0.4	115	0.00
51 T	4-Methyl-2-Pentanone	0.437	0.405	7.3	105	0.00
52 CM	Toluene	0.852	0.928	-8.9#	125	0.00
53 T	t-1,3-Dichloropropene	0.567	0.568	-0.2	115	0.00
54 T	cis-1,3-Dichloropropene	0.603	0.625	-3.6	120	0.00
55 T	1,1,2-Trichloroethane	0.337	0.366	-8.6	124	0.00
56 T	Ethyl methacrylate	0.522	0.528	-1.1	115	0.00
57 T	1,3-Dichloropropane	0.557	0.604	-8.4	123	0.00
58 T	2-Chloroethyl Vinyl ether	0.227	0.231	-1.8	118	0.00
59 T	2-Hexanone	0.321	0.306	4.7	106	0.00
60 T	Dibromochloromethane	0.402	0.417	-3.7	118	0.00
61 T	1,2-Dibromoethane	0.362	0.382	-5.5	120	0.00
62 S	4-Bromofluorobenzene	0.424	0.418	1.4	115	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	119	0.00
64 T	Tetrachloroethene	0.486	0.506	-4.1	127	0.00
65 PM	Chlorobenzene	1.018	1.121	-10.1	129	0.00
66 T	1,1,1,2-Tetrachloroethane	0.396	0.421	-6.3	124	0.00
67 C	Ethyl Benzene	1.834	1.925	-5.0#	123	0.00
68 T	m/p-Xylenes	0.691	0.738	-6.8	125	0.00
69 T	o-Xylene	0.671	0.721	-7.5	126	0.00
70 T	Styrene	1.119	1.241	-10.9	128	0.00
71 P	Bromoform	0.341	0.343	-0.6	116	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	121	0.00
73 T	Isopropylbenzene	3.808	3.714	2.5	122	0.00
74 T	N-amyl acetate	1.516	1.330	12.3	104	0.00
75 P	1,1,2,2-Tetrachloroethane	1.027	1.021	0.6	123	0.00
76 T	1,2,3-Trichloropropane	1.020	0.989	3.0	126	0.00
77 T	Bromobenzene	0.995	1.054	-5.9	131	0.00
78 T	n-propylbenzene	4.299	4.236	1.5	122	0.00
79 T	2-Chlorotoluene	2.532	2.519	0.5	122	0.00
80 T	1,3,5-Trimethylbenzene	3.247	3.186	1.9	122	0.00
81 T	trans-1,4-Dichloro-2-butene	0.405	0.361	10.9	108	0.00
82 T	4-Chlorotoluene	2.607	2.625	-0.7	124	0.00
83 T	tert-Butylbenzene	2.767	2.720	1.7	122	0.00
84 T	1,2,4-Trimethylbenzene	3.162	3.218	-1.8	125	0.00
85 T	sec-Butylbenzene	3.671	3.600	1.9	122	0.00
86 T	p-Isopropyltoluene	3.365	3.327	1.1	122	0.00
87 T	1,3-Dichlorobenzene	1.664	1.787	-7.4	130	0.00
88 T	1,4-Dichlorobenzene	1.671	1.776	-6.3	129	0.00
89 T	n-Butylbenzene	2.876	2.835	1.4	119	0.00
90 T	Hexachloroethane	0.579	0.557	3.8	114	0.00
91 T	1,2-Dichlorobenzene	1.598	1.767	-10.6	132	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.212	0.193	9.0	107	0.00
93 T	1,2,4-Trichlorobenzene	0.806	0.982	-21.8	126	0.00
94 T	Hexachlorobutadiene	0.631	0.591	6.3	120	0.00
95 T	Naphthalene	2.105	2.395	-13.8	120	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN010721\
Data File : VN065415.D
Acq On : 7 Jan 2021 9:56
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: Jan 08 01:13:44 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122220W.M
Quant Title : SW846 8260
QLast Update : Tue Dec 22 15:42:18 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.764	0.935	-22.4	133	0.02

(#) = Out of Range SPCC's out = 0 CCC's out = 6