

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN010725\
Data File : VN085395.D
Acq On : 07 Jan 2025 13:05
Operator : JC\MD
Sample : IBLK
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
IBLK

Quant Time: Jan 08 04:17:31 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N122024W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Sat Dec 21 02:03:07 2024
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	7.806	128	25198	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	9.100	114	114812	30.000	ug/l	0.00
57) Chlorobenzene-d5	11.865	117	107360	30.000	ug/l	0.00

System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	8.577	65	67351	30.384	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	101.267%
60) 4-Bromofluorobenzene	12.847	95	40514	23.296	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	77.667%
63) Toluene-d8	10.565	98	147012	28.043	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	93.467%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN010725\
Data File : VN085395.D
Acq On : 07 Jan 2025 13:05
Operator : JC\MD
Sample : IBLK
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
IBLK

Quant Time: Jan 08 04:17:31 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N122024W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Sat Dec 21 02:03:07 2024
Response via : Initial Calibration

