

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN010821\
 Data File : VN065439.D
 Acq On : 8 Jan 2021 13:16
 Operator : JC/MD
 Sample : VN0108WBL01
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0108WBL01

Quant Time: Jan 08 22:45:26 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122220W.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 22 15:42:18 2020
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.630	168	299769	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.553	114	493927	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.379	117	515445	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.318	152	202326	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.990	65	169538	50.10	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 100.20%	
35) Dibromofluoromethane	7.553	113	151887	50.97	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 101.94%	
50) Toluene-d8	10.061	98	602087	53.91	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 107.82%	
62) 4-Bromofluorobenzene	12.376	95	231279	52.12	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 104.24%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN010821\
Data File : VN065439.D
Acq On : 8 Jan 2021 13:16
Operator : JC/MD
Sample : VN0108WBL01
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0108WBL01

Quant Time: Jan 08 22:45:26 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122220W.M
Quant Title : SW846 8260
QLast Update : Tue Dec 22 15:42:18 2020
Response via : Initial Calibration

