

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN010821\
 Data File : VN065445.D
 Acq On : 8 Jan 2021 15:49
 Operator : JC/MD
 Sample : M1022-08DL 10X
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 16790DL

Quant Time: Jan 08 22:48:23 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122220W.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 22 15:42:18 2020
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.627	168	283480	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.550	114	467916	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.379	117	503037	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.318	152	217140	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.990	65	159715	49.91	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	99.82%		
35) Dibromofluoromethane	7.553	113	140964	49.93	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	99.86%		
50) Toluene-d8	10.058	98	579033	54.73	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	109.46%		
62) 4-Bromofluorobenzene	12.376	95	235699	55.68	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	111.36%		
Target Compounds					Qvalue	
16) Acetone	3.791	43	210604	152.63	ug/l	97
25) 2-Butanone	6.814	43	5599	3.12	ug/l	94
52) Toluene	10.129	92	1666	0.21	ug/l	91
68) m/p-Xylenes	11.592	106	1485	0.21	ug/l	82
95) Naphthalene	15.106	128	41919	6.85	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN010821\
Data File : VN065445.D
Acq On : 8 Jan 2021 15:49
Operator : JC/MD
Sample : M1022-08DL 10X
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
16790DL

Quant Time: Jan 08 22:48:23 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122220W.M
Quant Title : SW846 8260
QLast Update : Tue Dec 22 15:42:18 2020
Response via : Initial Calibration

