

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN010821\
 Data File : VN065448.D
 Acq On : 8 Jan 2021 17:03
 Operator : JC/MD
 Sample : PB133969TB
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB133969TB

Quant Time: Jan 08 22:49:29 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122220W.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 22 15:42:18 2020
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.630	168	289075	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.553	114	479303	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.380	117	522694	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.318	152	215029	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.994	65	164887	50.53	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery = 101.06%			
35) Dibromofluoromethane	7.557	113	143359	49.58	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery = 99.16%			
50) Toluene-d8	10.061	98	596466	55.04	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery = 110.08%			
62) 4-Bromofluorobenzene	12.376	95	237079	54.77	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery = 109.54%			
Target Compounds						
20) Methylene Chloride	4.518	84	8700	2.73	ug/l	94
43) Isopropyl Acetate	8.135	43	28286	3.80	ug/l	95
95) Naphthalene	15.109	128	11302	4.32	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN010821\
Data File : VN065448.D
Acq On : 8 Jan 2021 17:03
Operator : JC/MD
Sample : PB133969TB
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB133969TB

Quant Time: Jan 08 22:49:29 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122220W.M
Quant Title : SW846 8260
QLast Update : Tue Dec 22 15:42:18 2020
Response via : Initial Calibration

