

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN010821\
 Data File : VN065455.D
 Acq On : 8 Jan 2021 19:56
 Operator : JC/MD
 Sample : PB133969ZHE#07
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB133969ZHE#07

Quant Time: Jan 08 22:51:34 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122220W.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 22 15:42:18 2020
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.630	168	269712	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.553	114	456893	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.380	117	489433	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.318	152	194555	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.994	65	155701	51.14	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery = 102.28%			
35) Dibromofluoromethane	7.557	113	136950	49.68	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery = 99.36%			
50) Toluene-d8	10.061	98	565187	54.71	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery = 109.42%			
62) 4-Bromofluorobenzene	12.376	95	216672	52.72	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery = 105.44%			
Target Compounds						
20) Methylene Chloride	4.515	84	8660	2.92	ug/l	95
43) Isopropyl Acetate	8.135	43	25671	3.62	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN010821\
Data File : VN065455.D
Acq On : 8 Jan 2021 19:56
Operator : JC/MD
Sample : PB133969ZHE#07
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB133969ZHE#07

Quant Time: Jan 08 22:51:34 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122220W.M
Quant Title : SW846 8260
QLast Update : Tue Dec 22 15:42:18 2020
Response via : Initial Calibration

