

Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN010919\
 Data File : VN053319.D
 Acq On : 9 Jan 2019 9:07
 Operator : MD\SY
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 2 Sample Multiplier: 28

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Jan 10 03:22:48 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N010519W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jan 07 09:34:27 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	123	0.00
2 T	Dichlorodifluoromethane	0.441	0.438	0.7	119	0.00
3 P	Chloromethane	0.616	0.554	10.1	113	0.00
4 C	Vinyl Chloride	0.588	0.556	5.4#	116	0.00
5 T	Bromomethane	0.384	0.367	4.4	122	0.00
6 T	Chloroethane	0.355	0.333	6.2	120	0.00
7 T	Trichlorofluoromethane	0.761	0.735	3.4	122	0.00
8 T	Diethyl Ether	0.298	0.285	4.4	118	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.482	0.479	0.6	126	0.00
10 T	Methyl Iodide	0.676	0.689	-1.9	123	0.00
11 T	Tert butyl alcohol	0.031	0.032	-3.2	118	0.00
12 CM	1,1-Dichloroethene	0.467	0.447	4.3#	119	0.00
13 T	Acrolein	0.039	0.035	10.3	112	0.00
14 T	Allyl chloride	0.752	0.701	6.8	119	0.00
15 T	Acrylonitrile	0.187	0.176	5.9	114	0.00
16 T	Acetone	0.136	0.159	-16.9	155#	0.00
17 T	Carbon Disulfide	1.438	1.337	7.0	117	0.00
18 T	Methyl Acetate	0.471	0.408	13.4	116	0.00
19 T	Methyl tert-butyl Ether	1.309	1.260	3.7	119	0.00
20 T	Methylene Chloride	0.559	0.509	8.9	118	0.00
21 T	trans-1,2-Dichloroethene	0.501	0.481	4.0	120	0.00
22 T	Diisopropyl ether	1.620	1.559	3.8	118	0.00
23 T	Vinyl Acetate	1.019	0.989	2.9	117	0.00
24 P	1,1-Dichloroethane	0.934	0.888	4.9	118	0.00
25 T	2-Butanone	0.213	0.214	-0.5	124	0.00
26 T	2,2-Dichloropropane	0.683	0.689	-0.9	127	0.00
27 T	cis-1,2-Dichloroethene	0.572	0.554	3.1	119	0.00
28 T	Bromochloromethane	0.398	0.388	2.5	116	0.00
29 T	Tetrahydrofuran	0.141	0.131	7.1	114	0.00
30 C	Chloroform	0.920	0.873	5.1#	118	0.00
31 T	Cyclohexane	1.024	0.834	18.6	118	0.00
32 T	1,1,1-Trichloroethane	0.766	0.730	4.7	119	0.00
33 S	1,2-Dichloroethane-d4	0.520	0.493	5.2	118	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	121	0.00
35 S	Dibromofluoromethane	0.285	0.286	-0.4	122	0.00
36 T	1,1-Dichloropropene	0.467	0.466	0.2	122	0.00
37 T	Ethyl Acetate	0.304	0.293	3.6	115	0.00
38 T	Carbon Tetrachloride	0.431	0.430	0.2	120	0.00
39 T	Methylcyclohexane	0.541	0.543	-0.4	121	0.00
40 TM	Benzene	1.426	1.400	1.8	119	0.00
41 T	Methacrylonitrile	0.189	0.150	20.6	103	-0.01
42 TM	1,2-Dichloroethane	0.423	0.408	3.5	118	0.00
43 T	Isopropyl Acetate	0.598	0.526	12.0	114	0.00
44 TM	Trichloroethene	0.392	0.388	1.0	121	0.00
45 C	1,2-Dichloropropane	0.374	0.369	1.3#	119	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN010919\
 Data File : VN053319.D
 Acq On : 9 Jan 2019 9:07
 Operator : MD\SY
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 2 Sample Multiplier: 28

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Jan 10 03:22:48 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N010519W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jan 07 09:34:27 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.218	0.214	1.8	117	0.00
47 T	Bromodichloromethane	0.454	0.452	0.4	120	0.00
48 T	Methyl methacrylate	0.272	0.263	3.3	114	0.00
49 T	1,4-Dioxane	0.003	0.003	0.0	123	0.00
50 S	Toluene-d8	1.228	1.215	1.1	120	0.00
51 T	4-Methyl-2-Pentanone	0.301	0.294	2.3	115	0.00
52 CM	Toluene	0.875	0.877	-0.2#	119	0.00
53 T	t-1,3-Dichloropropene	0.465	0.478	-2.8	121	0.00
54 T	cis-1,3-Dichloropropene	0.537	0.549	-2.2	120	0.00
55 T	1,1,2-Trichloroethane	0.322	0.311	3.4	118	0.00
56 T	Ethyl methacrylate	0.418	0.427	-2.2	118	0.00
57 T	1,3-Dichloropropane	0.538	0.532	1.1	119	0.00
58 T	2-Chloroethyl Vinyl ether	0.226	0.222	1.8	115	0.00
59 T	2-Hexanone	0.206	0.213	-3.4	120	0.00
60 T	Dibromochloromethane	0.347	0.351	-1.2	119	0.00
61 T	1,2-Dibromoethane	0.317	0.311	1.9	116	0.00
62 S	4-Bromofluorobenzene	0.430	0.428	0.5	122	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	121	0.00
64 T	Tetrachloroethene	0.471	0.465	1.3	118	0.00
65 PM	Chlorobenzene	1.086	1.070	1.5	119	0.00
66 T	1,1,1,2-Tetrachloroethane	0.380	0.379	0.3	120	0.00
67 C	Ethyl Benzene	1.846	1.870	-1.3#	120	0.00
68 T	m/p-Xylenes	0.694	0.710	-2.3	120	0.00
69 T	o-Xylene	0.679	0.686	-1.0	119	0.00
70 T	Styrene	1.106	1.140	-3.1	120	0.00
71 P	Bromoform	0.269	0.274	-1.9	120	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	117	0.00
73 T	Isopropylbenzene	3.723	3.842	-3.2	119	0.00
74 T	N-amyl acetate	1.112	1.124	-1.1	114	0.00
75 P	1,1,2,2-Tetrachloroethane	0.856	0.836	2.3	117	0.00
76 T	1,2,3-Trichloropropane	0.729	0.748	-2.6	124	0.00
77 T	Bromobenzene	0.994	1.008	-1.4	120	0.00
78 T	n-propylbenzene	4.211	4.387	-4.2	119	0.00
79 T	2-Chlorotoluene	2.548	2.571	-0.9	118	0.00
80 T	1,3,5-Trimethylbenzene	2.999	3.085	-2.9	119	0.00
81 T	trans-1,4-Dichloro-2-butene	0.248	0.276	-11.3	122	0.00
82 T	4-Chlorotoluene	2.613	2.665	-2.0	120	0.00
83 T	tert-Butylbenzene	2.631	2.664	-1.3	119	0.00
84 T	1,2,4-Trimethylbenzene	3.055	3.143	-2.9	119	0.00
85 T	sec-Butylbenzene	3.470	3.560	-2.6	119	0.00
86 T	p-Isopropyltoluene	3.004	3.125	-4.0	119	0.00
87 T	1,3-Dichlorobenzene	1.750	1.744	0.3	118	0.00
88 T	1,4-Dichlorobenzene	1.736	1.731	0.3	119	0.00
89 T	n-Butylbenzene	2.519	2.626	-4.2	121	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN010919\
Data File : VN053319.D
Acq On : 9 Jan 2019 9:07
Operator : MD\SY
Sample : VSTDCCC050
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 2 Sample Multiplier: 28

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: Jan 10 03:22:48 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N010519W.M
Quant Title : SW846 8260
QLast Update : Mon Jan 07 09:34:27 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.481	0.493	-2.5	118	0.00
91 T	1,2-Dichlorobenzene	1.639	1.624	0.9	116	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.104	0.109	-4.8	116	0.00
93 T	1,2,4-Trichlorobenzene	0.515	0.560	-8.7	121	0.00
94 T	Hexachlorobutadiene	0.421	0.445	-5.7	121	0.00
95 T	Naphthalene	0.879	1.002	-14.0	125	0.00
96 T	1,2,3-Trichlorobenzene	0.337	0.374	-11.0	125	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6