

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN010923\
 Data File : VN076146.D
 Acq On : 09 Jan 2023 11:06
 Operator : JC\MD
 Sample : VN0109MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0109MBL01

Quant Time: Jan 10 01:02:03 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121422W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 14 13:03:41 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.236	168	262874	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	413571	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	357749	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	147380	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.589	65	143313	50.113	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	100.220%
35) Dibromofluoromethane	8.177	113	127005	49.588	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.180%
50) Toluene-d8	10.571	98	476292	49.738	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	99.480%
62) 4-Bromofluorobenzene	12.853	95	145186	45.776	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	91.560%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN010923\
Data File : VN076146.D
Acq On : 09 Jan 2023 11:06
Operator : JC\MD
Sample : VN0109MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0109MBL01

Quant Time: Jan 10 01:02:03 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121422W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 14 13:03:41 2022
Response via : Initial Calibration

