

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN010923\
 Data File : VN076156.D
 Acq On : 09 Jan 2023 15:16
 Operator : JC\MD
 Sample : PB150106ZHE#01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB150106ZHE#01

Quant Time: Jan 10 01:07:22 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121422W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 14 13:03:41 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.229	168	325561	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	529513	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	481661	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	208139	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	175498	49.551	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	99.100%
35) Dibromofluoromethane	8.177	113	158678	48.389	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.780%
50) Toluene-d8	10.570	98	631951	51.543	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	103.080%
62) 4-Bromofluorobenzene	12.853	95	205470	50.598	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	101.200%
Target Compounds						
					Qvalue	
16) Acetone	4.441	43	37339	30.582	ug/l	90
20) Methylene Chloride	5.294	84	11673	2.972	ug/l	88
25) 2-Butanone	7.494	43	9000	5.265	ug/l	98
37) Ethyl Acetate	7.571	43	31790	9.085	ug/l	98
43) Isopropyl Acetate	8.694	43	170446	29.531	ug/l #	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN010923\
Data File : VN076156.D
Acq On : 09 Jan 2023 15:16
Operator : JC\MD
Sample : PB150106ZHE#01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB150106ZHE#01

Quant Time: Jan 10 01:07:22 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121422W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 14 13:03:41 2022
Response via : Initial Calibration

