

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN010923\
 Data File : VN076157.D
 Acq On : 09 Jan 2023 15:40
 Operator : JC\MD
 Sample : PB150106ZHE#02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB150106ZHE#02

Quant Time: Jan 10 01:07:51 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121422W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 14 13:03:41 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	304648	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	498282	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	452467	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	185743	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	165442	49.918	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	99.840%		
35) Dibromofluoromethane	8.171	113	150455	48.757	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	97.520%		
50) Toluene-d8	10.571	98	594394	51.519	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	103.040%		
62) 4-Bromofluorobenzene	12.853	95	187979	49.192	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	98.380%		
Target Compounds					Qvalue	
16) Acetone	4.442	43	36700	32.123	ug/l	99
20) Methylene Chloride	5.294	84	10994	2.995	ug/l #	92
25) 2-Butanone	7.506	43	8472	5.296	ug/l	97
37) Ethyl Acetate	7.571	43	31855	9.675	ug/l	99
43) Isopropyl Acetate	8.694	43	226930	41.781	ug/l #	76

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN010923\
Data File : VN076157.D
Acq On : 09 Jan 2023 15:40
Operator : JC\MD
Sample : PB150106ZHE#02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB150106ZHE#02

Quant Time: Jan 10 01:07:51 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121422W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 14 13:03:41 2022
Response via : Initial Calibration

