

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN010923\
 Data File : VN076160.D
 Acq On : 09 Jan 2023 16:52
 Operator : JC\MD
 Sample : 01065-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VNJ-210

Quant Time: Jan 10 01:09:21 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121422W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 14 13:03:41 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	311242	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	518071	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	458797	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	184990	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	172535	50.955	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	101.920%
35) Dibromofluoromethane	8.171	113	151800	47.314	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.620%
50) Toluene-d8	10.571	98	599011	49.936	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	99.880%
62) 4-Bromofluorobenzene	12.853	95	190242	47.883	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	95.760%
Target Compounds						
						Qvalue
16) Acetone	4.448	43	49542	42.444	ug/l	97
18) Methyl Acetate	5.047	43	3578	1.037	ug/l	91
20) Methylene Chloride	5.295	84	9427	2.410	ug/l #	82
25) 2-Butanone	7.494	43	8870	5.427	ug/l	99
37) Ethyl Acetate	7.571	43	23773	6.944	ug/l	97
43) Isopropyl Acetate	8.694	43	291700	51.654	ug/l #	68

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN010923\
Data File : VN076160.D
Acq On : 09 Jan 2023 16:52
Operator : JC\MD
Sample : 01065-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VNJ-210

Quant Time: Jan 10 01:09:21 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121422W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 14 13:03:41 2022
Response via : Initial Calibration

